

**बिड दस्तावेज़ / Bid Document**

बिड विवरण/Bid Details	
बिड बंद होने की तारीख/समय /Bid End Date/Time	07-02-2026 15:00:00
बिड खुलने की तारीख/समय /Bid Opening Date/Time	07-02-2026 15:30:00
बिड पेशकश वैधता (बंद होने की तारीख से)/Bid Offer Validity (From End Date)	180 (Days)
मंत्रालय/राज्य का नाम/Ministry/State Name	Ministry Of Health And Family Welfare
विभाग का नाम/Department Name	Department Of Health And Family Welfare
संगठन का नाम/Organisation Name	All India Institute Of Medical Sciences (aiims)
कार्यालय का नाम/Office Name	Guwahati
कुल मात्रा/Total Quantity	11
वस्तु श्रेणी /Item Category	ICU Ventilators (V2) (Q2)
बिडर का न्यूनतम औसत वार्षिक टर्नओवर (3 वर्षों का) /Minimum Average Annual Turnover of the bidder (For 3 Years)	176 Lakh (s)
मूल उपकरण निर्माता का औसत टर्नओवर (गत 3 वर्षों का)/OEM Average Turnover (Last 3 Years)	1408 Lakh (s)
उन्हीं/समान सेवा के लिए अपेक्षित विगत अनुभव के वर्ष/Years of Past Experience Required for same/similar service	3 Year (s)
वर्षों के अनुभव एवं टर्नओवर से एमएसई को छूट प्राप्त है / MSE Relaxation for Years Of Experience and Turnover	Yes   Complete
स्टार्टअप के लिए अनुभव के वर्षों और टर्नओवर से छूट प्रदान की गई है /Startup Relaxation for Years of Experience and Turnover	No
विक्रेता से मांगे गए दस्तावेज़/Document required from seller	Experience Criteria,Past Performance,Bidder Turnover,Certificate (Requested in ATC),OEM Authorization Certificate,OEM Annual Turnover,Additional Doc 1 (Requested in ATC),Additional Doc 2 (Requested in ATC) *In case any bidder is seeking exemption from Experience / Turnover Criteria, the supporting documents to prove his eligibility for exemption must be uploaded for evaluation by the buyer

**बिड विवरण/Bid Details**

क्या आप निविदाकारों द्वारा अपलोड किए गए दस्तावेजों को निविदा में भाग लेने वाले सभी निविदाकारों को दिखाना चाहते हैं? संदर्भ मेनू है/ <b>Do you want to show documents uploaded by bidders to all bidders participated in bid?</b>	Yes (Documents submitted as part of a clarification or representation during the tender/bid process will also be displayed to other participated bidders after log in)
बिड लगाने की समय सीमा स्वतः नहीं बढ़ाने के लिए आवश्यक बिड की संख्या। / <b>Minimum number of bids required to disable automatic bid extension</b>	3
दिनों की संख्या, जिनके लिए बिड लगाने की समय-सीमा बढ़ाई जाएगी। / <b>Number of days for which Bid would be auto-extended</b>	7
ऑटो एक्सटेंशन अधिकतम कितनी बार किया जाना है। / <b>Number of Auto Extension count</b>	2
विगत प्रदर्शन / <b>Past Performance</b>	10 %
बिड से रिवर्स नीलामी सक्रिय किया/ <b>Bid to RA enabled</b>	No
व्यापक रखरखाव शुल्क आवश्यक / <b>Comprehensive Maintenance Charges Required</b>	Yes
बिड का प्रकार/ <b>Type of Bid</b>	Two Packet Bid
तकनीकी मूल्यांकन के दौरान तकनीकी स्पष्टीकरण हेतु अनुमत समय / <b>Time allowed for Technical Clarifications during technical evaluation</b>	2 Days
निरीक्षण आवश्यक (सूचीबद्ध निरीक्षण प्राधिकरण /जेम के साथ पूर्व पंजीकृत एजेंसियों द्वारा)/ <b>Inspection Required (By Empanelled Inspection Authority / Agencies pre-registered with GeM)</b>	No
मूल्यांकन पद्धति/ <b>Evaluation Method</b>	Total value wise evaluation
वित्तीय दस्तावेज की आवश्यकता है / <b>Financial Document Required</b>	Yes
मध्यस्थता खंड/ <b>Arbitration Clause</b>	No
सुलह खंड/ <b>Mediation Clause</b>	No

**ईएमडी विवरण/EMD Detail**

एडवाइजरी बैंक/ <b>Advisory Bank</b>	ICICI
ईएमडी राशि/ <b>EMD Amount</b>	704000

**ईपीबीजी विवरण /ePBG Detail**

--	--

एडवाइजरी बैंक/Advisory Bank	ICICI
ईपीबीजी प्रतिशत (%) / ePBG Percentage (%)	5.00
ईपीबीजी की आवश्यक अवधि (माह) / Duration of ePBG required (Months).	26

(a). जेम की शर्तों के अनुसार ईएमडी छूट के इच्छुक बिडर को संबंधित केटेगरी के लिए बिड के साथ वैध समर्थित दस्तावेज प्रस्तुत करने है। एमएसई केटेगरी के अंतर्गत केवल वस्तुओं के लिए विनिर्माता तथा सेवाओं के लिए सेवा प्रदाता ईएमडी से छूट के पात्र हैं। व्यापारियों को इस नीति के दायरे से बाहर रखा गया है।/EMD EXEMPTION: The bidder seeking EMD exemption, must submit the valid supporting document for the relevant category as per GeM GTC with the bid. Under MSE category, only manufacturers for goods and Service Providers for Services are eligible for exemption from EMD. Traders are excluded from the purview of this Policy.

(b). ईएमडी और संपादन जमानत राशि, जहां यह लागू होती है, लाभार्थी के पक्ष में होनी चाहिए। / EMD & Performance security should be in favour of Beneficiary, wherever it is applicable.

**लाभार्थी /Beneficiary :**

AIIMS GUWAHATI EMD/SD  
Guwahati, Department of Health and Family Welfare, All India Institute of Medical Sciences (AIIMS), Ministry of Health and Family Welfare  
(Aaims/guwahati Emd/sd)

बोली विभाजन लागू नहीं किया गया/ Bid splitting not applied.

**एमआईआई खरीद वरीयता/MII Purchase Preference**

एमआईआई खरीद वरीयता/MII Purchase Preference	No
--	----

**एमआईआई के लिए सक्षम प्राधिकारी का विवरण:/Details of the Competent Authority for MII**

सक्षम प्राधिकारी का नाम/Name of Competent Authority	Sher Bahadur
सक्षम प्राधिकारी का पदनाम/Designation of Competent Authority	Under Secretary (Procurement Policy)
सक्षम प्राधिकारी का कार्यालय/विभाग/प्रभाग/Office / Department / Division of Competent Authority	Government of India Ministry of Finance Department of Expenditure Procurement Policy Division
सीए अनुमोदन संख्या/CA Approval Number	No.F .41 1 t2023-PPD(pt.)
सक्षम प्राधिकारी अनुमोदन तिथि/Competent Authority Approval Date	17-02-2025
सक्षम प्राधिकारी द्वारा प्रदान की गई स्वीकृति का संक्षिप्त विवरण/Brief Description of the Approval Granted by Competent Authority	Relaxation under Rule 161(iv) of General Financial Rules 2017 for issuance of Global Tender Enquiry (GTE) for procurement of Medical Devices -reg.

Competent Authority Approval for not opting Make In India Preference : [View Document](#)

**एमएसई खरीद वरीयता/MSE Purchase Preference**

एमएसई खरीद वरीयता/MSE Purchase Preference	Yes
---	-----

सूक्ष्म और लघु उद्यम मूल उपकरण निर्माताओं को खरीद में प्राथमिकता, यदि उनका मूल्य L1+X% तक की सीमा में हो / Purchase Preference to MSE OEMs available upto price within L1+X%	15
सूक्ष्म और लघु उद्यम को खरीद में प्राथमिकता के लिए बिड की मात्रा का अधिकतम प्रतिशत / Maximum Percentage of Bid quantity for MSE purchase preference	25

1. If the bidder is a Micro or Small Enterprise as per latest orders issued by Ministry of MSME, the bidder shall be relaxed from the eligibility criteria of "Experience Criteria" as defined above subject to meeting of quality and technical specifications. The bidder seeking Relaxation from Experience Criteria, shall upload the supporting documents to prove his eligibility for Relaxation.
2. If the bidder is a Micro or Small Enterprise (MSE) as per latest orders issued by Ministry of MSME, the bidder shall be relaxed from the eligibility criteria of "Bidder Turnover" as defined above subject to meeting of quality and technical specifications. If the bidder itself is MSE OEM of the offered products, it would be relaxed from the "OEM Average Turnover" criteria also subject to meeting of quality and technical specifications. The bidder seeking Relaxation from Turnover, shall upload the supporting documents to prove his eligibility for Relaxation.
3. The minimum average annual financial turnover of the bidder during the last three years, ending on 31st March of the previous financial year, should be as indicated above in the bid document. Documentary evidence in the form of certified Audited Balance Sheets of relevant periods or a certificate from the Chartered Accountant / Cost Accountant indicating the turnover details for the relevant period shall be uploaded with the bid. In case the date of constitution / incorporation of the bidder is less than 3-year-old, the average turnover in respect of the completed financial years after the date of constitution shall be taken into account for this criteria.
4. Experience Criteria: In respect of the filter applied for experience criteria, the Bidder or its OEM of the product offered in the bid {themselves or through reseller(s)} should have regularly, manufactured and supplied same or similar Category Products to any Central / State Govt Organization / PSU for number of Financial years as indicated above in the bid document before the bid opening date. Copies of relevant contracts and delivery acceptance certificates like CRAC to be submitted along with bid in support of having supplied some quantity during each of the Financial year. In case of bunch bids, the category of primary product having highest value should meet this criterion.
5. OEM Turn Over Criteria: The minimum average annual financial turnover of the OEM of the offered product during the last three years, ending on 31st March of the previous financial year, should be as indicated in the bid document. Documentary evidence in the form of certified Audited Balance Sheets of relevant periods or a certificate from the Chartered Accountant / Cost Accountant indicating the turnover details for the relevant period shall be uploaded with the bid. In case the date of constitution / incorporation of the OEM is less than 3 year old, the average turnover in respect of the completed financial years after the date of constitution shall be taken into account for this criteria.
6. Purchase preference will be given to MSEs having valid Udyam Registration and whose credentials are validated online through Udyam Registration portal as defined in Public Procurement Policy for Micro and Small Enterprises (MSEs) Order, 2012 dated 23.03.2012 issued by Ministry of Micro, Small and Medium Enterprises and its subsequent Orders/Notifications issued by concerned Ministry. If the bidder wants to avail themselves of the Purchase preference, the bidder must be the manufacturer / OEM of the offered product on GeM. Traders are excluded from the purview of Public Procurement Policy for Micro and Small Enterprises and hence resellers offering products manufactured by some other OEM are not eligible for any purchase preference. In respect of bid for Services, the bidder must be the Service provider of the offered Service. Relevant documentary evidence in this regard shall be uploaded along with the bid in respect of the offered product or service and Buyer will decide eligibility for purchase preference based on documentary evidence submitted, while evaluating the bid. If L-1 is not an MSE and MSE Seller (s) has / have quoted price within L-1+ 15% (Selected by Buyer) of margin of purchase preference /price band defined in relevant policy, such MSE Seller shall be given opportunity to match L-1 price and contract will be awarded for 25% (selected by Buyer) percentage of total quantity. The buyers are advised to refer the OM No. F.1/4/2021-PPD dated 18.05.2023 [OM No.1 4 2021 PPD dated 18.05.2023](#) for compliance of Concurrent application of Public Procurement Policy for Micro and Small Enterprises Order, 2012 and Public Procurement (Preference to Make in India) Order, 2017. Benefits of MSE will be allowed only if seller is validated on-line in GeM profile as well as validated and approved by Buyer after evaluation of documents submitted.
7. Estimated Bid Value indicated above is being declared solely for the purpose of guidance on EMD amount and for determining the Eligibility Criteria related to Turn Over, Past Performance and Project / Past Experience etc. This has no relevance or bearing on the price to be quoted by the bidders and is also not going to have any

impact on bid participation. Also this is not going to be used as a criteria in determining reasonableness of quoted prices which would be determined by the buyer based on its own assessment of reasonableness and based on competitive prices received in Bid / RA process.

8. Past Performance: The Bidder or its OEM {themselves or through re-seller(s)} should have supplied same or similar Category Products for 10% of bid quantity, in at least one of the last three Financial years before the bid opening date to any Central / State Govt Organization / PSU. Copies of relevant contracts (proving supply of cumulative order quantity in any one financial year) to be submitted along with bid in support of quantity supplied in the relevant Financial year. In case of bunch bids, the category related to primary product having highest bid value should meet this criterion.

## ICU Ventilators (V2) ( 11 pieces )

### तकनीकी विशिष्टियाँ /Technical Specifications

\* जेम केटेगरी विशिष्टि के अनुसार / As per GeM Category Specification

विवरण/Specification	विशिष्टि का नाम /Specification Name	बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)
GENERAL	Product Description	ICU Ventilator
PRODUCT INFORMATION	<b>Patient Type</b>	Adult & pediatric
	<b>Type of technology</b>	Turbine
	<b>Tidal volume in ml</b>	(20 ml or less) to (1500 ml or more)
	<b>Respiration rate, breaths/min</b>	up to 80 or more
	<b>Trigger mechanism</b>	Pressure /flow
	FiO2%	21 to 100
	<b>Inspiratory flow rate, L/min</b>	upto 200 or more
	<b>Inspiratory pressure, cm H2O</b>	upto 50 or more
	<b>IE ratio</b>	1:10 to 4:1
	<b>Sigh breath function</b>	Yes
	PEEP/CPAP, cm H2O	upto 40 or more
	Pressure support, cm H2O	0 to 40 or more
	Leak compensation	Yes
	Auto 100%/Increase O2 button	Yes
	<b>Control panel lock</b>	Yes
<b>Facility for double lung ventilation</b>	Yes	
PATIENT ASSESSMENT TOOLS	Maximum wave forms displayed	3 (Pressure and time, volume and time , flow and time)
	<b>Number of loops</b>	2 loops (P-V, F-V)
	Maximum trending time in hrs	?24 hr
	<b>Lung recruitment tools (PV loops) / OLT (Open lung tools)</b>	No

विवरण/Specification	विशिष्टि का नाम /Specification Name	बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)
	<b>Lung mechanics visualization tool</b>	No
	<b>Capnography/CO2 monitoring</b>	Yes
	<b>Esophageal / transpulmonary pressure monitoring</b>	No
	Modes of ventilation	Volume controlled, Pressure Controlled, Pressure Support, SIMV with Pressure support, SIMV with volume control with pressure support, CPAP/PEEP, Inverse Ratio Ventilation, Non invasive ventilation, Apnea /back up ventilation, SIMV (VC) with Pressure support, SIMV (PC) with Pressure Support, SIMV (PRVC) with Pressure Support
	<b>APRV(Airway Pressure Release Ventilation) /Bivent/Bi level</b>	Yes
	<b>MMV(Mandatory Minute Ventilation) + PSV (Pressure Support Ventilation) /ASV( Adaptive Support Ventilation ) or ALPV(Adaptive Lung Protection Ventilation) or Automode</b>	Yes
	<b>PRVC(Pressure-Regulated Volume Control ventilation)/Auto flow</b>	Yes
	<b>High flow oxygen therapy</b>	Yes
	<b>Automatic weaning system provided</b>	Yes
	<b>Volume Support</b>	No
	Ventilation monitoring facility	The Ventilator is a Microprocessor Controlled ventilator with integrated facility for Ventilation monitoring
	Monitored / Displayed parameters	Peak inspiratory pressure, Mean airway pressure, PEEP pressure, Tidal volume, Minute volume, Spontaneous minute volume, FiO2 (analyzed %), Respiratory rate, Inspiratory time, Expiratory time, IE ratio, Plateau Pressure
	Availability of Status indicator on display	Ventilator mode, Battery life, patient data, alarm settings, clock
	Patient alarms	Low/high FiO2, Low/high minute volume, Low inspiratory pressure, High pressure, Low PEEP, High PEEP, Apnea, Continuous high pressure/occlusion, Inverse IE, High respiratory rate, Breathing circuit disconnect
	Equipment alarms	Gas-supply failure, Power failure, Vent inoperative, Low battery, Self-diagnostic

विवरण/Specification	विशिष्टि का नाम /Specification Name	बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)
	<b>RSBI (Rapid Shallow Breathing Index)</b>	Yes
	<b>Expiratory Time constant ( Tcexp)</b>	Yes
	<b>Compliance static and/or dynamic</b>	Yes
	<b>Inspired and/or expired resistance</b>	Yes
	<b>Occlusion pressure</b>	No
	<b>Inspiratory and expiratory hold</b>	Yes
	<b>Spontaneous frequency</b>	Yes
	<b>Total peep, intrinsic peep, extrinsic peep</b>	Yes
	<b>Expiratory block is autoclavable and no routine calibration required</b>	Yes
	<b>Auto compensation for ET Tube</b>	No
	<b>Facility for automatic compliance &amp; Leakage compensation for circuit available</b>	Yes
	<b>RS 232 out put port</b>	Yes
	<b>USB/Memory card</b>	No
	<b>Availability of Remote alarm/display port</b>	Yes
	<b>Facility to Report (vent alarms and patient status)</b>	Yes
	<b>HL/7 Compliant out put</b>	Yes
	<b>Type of coloured display</b>	LED/LCD/TFT
	<b>Display should be touch screen</b>	Yes
	<b>NIV (Non Invasive Ventilation) to be possible in all modes of ventilation available</b>	Yes
	<b>Minimum Size of display (±2%)</b>	12.0 (inch)
	<b>Graphic display have automatic scaling facility for waves</b>	Yes

विवरण/Specification	विशिष्टि का नाम /Specification Name	बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)
HARDWARE AND ELECTRICAL REQUIREMENTS	Power supply	220-240 V , 50 Hz AC
	<b>Provision of UPS</b>	No
	<b>Backup time in hrs</b>	NA if not provided
	Internal backup battery	Yes
	Backup time of internal battery in mins	30 mins or more
	<b>Built in air source</b>	Turbine (In built)
	<b>Availability of stand alone compressor</b>	No
	<b>Compressor shall provide an oil free Medical air, with Peak output flow should be minimum 150 LPM and Air quality complying with ISO compressed air purity class</b>	NA (for turbine)
	Medical Air Compressor should automatically activate in the event of wall air supply loss and replacement of internal filters should be performed without removing the compressor, to have washable air filter	NA (for turbine)
	<b>Seller shall ensure compatibility of compressor with ventilator</b>	NA (for Turbine)
	<b>Nebulizer with particle size less than 3 micron</b>	Yes
ADDITIONAL ACCESSORIES	Number of reusable silicon breathing circuits	2
	NV mask(Re usable)- Small,medium and large	2 Each
	Length of Power cord in m	? 3 m
	Flow sensor Placement	Proximal
	Number of flow sensor	20 Disposable
	Autoclavable exhalation valve/expiratory cassette / filter	NA for flow Sensor
	<b>Humidifier: Servo controlled with digital monitoring of inspired gas temperature with heating wire</b>	1
	Trolley for ventilator with circuit holding arm	Yes
CERTIFICATIONS	Compliance to Medical Device Rules (MDR) 2017 as amended till date	Yes

विवरण/Specification	विशिष्टि का नाम /Specification Name	बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)
	Availability of valid Medical Device license for the product issued from the competent authority defined under Drugs and Cosmetic Act 1940 and Rules made there under as amended till date	Yes
	Certification for manufacturing unit	ISO:13485 (Latest)
	Availability of Test Report for each supplied batch/product as per Medical Device Rule (MDR) 2017 as amended till date	Yes
	Submission of all necessary certifications, licenses and test reports to the buyer at the time of bid submission or along with supplies as per buyer requirement	Yes
	Electrical Safety Standard	IEC 60601-1-2 or Equivalent BIS
	Safety and Essential Performance	IS 13450 Part 2 Section 12
MISCELLANEOUS PARAMETERS	OEM/Reseller (if supplied by reseller shall ensure uninterrupted availability of all spares for 10 years)	Yes
	Availability of toll free facility for technical support maintained by OEM or authorized agencies	Yes
	During the warranty period it shall be ensured that all the break down calls are attended with in 24 hrs and the complete details of service agents / contact details should be furnished to buyer and consignee at time of supplies	Yes
	User/Technical/Maintenance manuals to be supplied in English in hard and soft copy	Yes
	Details of equipments and procedures required for local calibration and routine maintenance to be supplied and advanced maintenance task documentation also to be furnished	Yes

विवरण/Specification	विशिष्टि का नाम /Specification Name	बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)
	List of important spares and accessories, with their part numbers and price lists to be supplied to the buyer at the time of supplying the equipment	Yes
	Demonstration of equipment and training to be provided after completing supplies before acceptance	Yes
	The Principal Manufacturer must have direct Presence/approved service center In India	Yes
	Copies of reports and certifications to be furnished to buyer on demand at time of supplies	Yes
WARRANTY	<b>Warranty (years)</b>	5 Or higher <b>(year)</b>

व्यापक रखरखाव / Comprehensive Maintenance	
Warranty of required product	2 Year
Comprehensive Maintenance Duration (Post Warranty)	8 Year

\*Warranty displayed under the AMC/CMC Details section will supersede the warranty displayed under the catalog specification

**परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity**

क्र.सं./S.N o.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Vineet Mahajan	781101,Changsari	11	45

**Special terms and conditions-Version:1 effective from 15-04-2024 for category ICU Ventilators (V2)**

1. All Provisions of Drugs and Cosmetics Act, 1940 and Rules (including Medical Device Rule 2017) made there under as amended till date will always be applicable. This will include all notifications issued by Central Drugs Standard Control Organization (CDSCO), Ministry of Health & Family Welfare (MoHFW) and Department of Pharmaceuticals (DOP), Ministry of Chemicals & Fertilizers time to time in this regard.
2. The sellers are registered on GeM based on the self declaration of valid Medical Device License,

- product certification, test reports etc. However, buyers must check and validate the details at their end for all applicable licenses and certifications e.g., validity and authenticity/genuineness of Medical Device license, product certification, manufacturer certification/licenses, test reports etc.
3. In case of authorized resellers/distributors, it will be the legal & regulatory liability of the manufacturer to ensure that their resellers/distributors are operating in compliance with all relevant laws and regulations and are properly licensed to sell the manufacturer's products, including verifying the validity and authenticity of Medical Device license held by them.
  4. The price offered by the seller/bidder shall not, in any case exceed the DPCO/NPPA controlled price or price fixed by State Government, if any. The seller must reduce the prices if there is any reduction in DPCO/NPPA ceiling price or price fixed by State Government, if any.
  5. Any other Terms and Conditions which is not included or at variance with the conditions specified in STC/GTC, may be added by the buyer through Additional Terms and Conditions (ATC) in the bid to ensure items are procured from authentic/validated source with appropriate and applicable quality. The above terms and conditions are in reverse order of precedence i.e. ATC shall supersede specific STC which shall supersede General Terms and Conditions (GTC), whenever there are any conflicting provisions.
  6. **Comprehensive warranty:** Comprehensive warranty shall include preventive maintenance including calibration as per technical/ service /operational manual of the manufacturer, service charges and spares. During the warranty period commencing from date of the successful completion of warranty period, Service personnel shall visit each consignee site as recommended in the manufacturer's technical/ service /operational manual, at least once in six months. warranty shall not be including the consumables. Further there will be 98% uptime warranty during warranty period on 24 (hrs) X 7 (days) X 365 (days) basis, with penalty, to extend warranty period by double the downtime period.
  7. **Service centres:** Details of Service outlets in India to render services for equipment to be furnished to buyer/consignees with complete address, telephone numbers, e mails etc at time of making the supplies. It shall be the responsibility of seller to ensure that authorized service centres are available to cater to the areas where supplies are made within reasonable distance from where the service calls can be handled. Details of toll-free numbers for service call and online registration of service requests also to be provided buyer/consignee at the time of supplies.
  8. **Source of supply:** It shall be responsibility of seller to provide Documents regarding source of equipments such as copy of Performa invoice or any other documents to establish that the products supplied are manufactured by OEM indicated and sourced from them.
  9. **Packing and Marking:** Medical equipments being very delicate and sensitive packing for the goods should be strong and durable enough to withstand transit including transshipment (if any), rough handling, open storage etc. without any damage, deterioration etc. .The size, weights and volumes of the packing cases, remoteness of the final destination of the goods, availability or otherwise of transport and handling facilities at all points during transit up to final destination,. Quality of packing, the manner of marking within & outside the packages and provision of accompanying documentation shall take into consideration the type of medical equipments being supplied. The accessories shall be suitably labelled and packed. Each of the package shall be marked on three sides with indelible paint of proper quality: indicating contract number and date, brief description of goods including quantity, Packing list reference number, country of origin of goods and any other relevant details.
  10. **Spare Parts:** Seller shall provide materials, information etc. pertaining to spare parts manufactured and supplied by the OEM. It shall be ensured that the required spares are available for purchase at least for 10 years from date of supplies. In case due to any reasons the production of the spare parts is discontinued sufficient advance notice should be given to the buyer/consignee before such discontinuation to provide adequate time to purchase the required spare parts etc. Further, OEM and their service centres/dealers shall carry sufficient inventories to assure ex-stock supply of consumables and spares for the equipments so that the same are available. OEM or reseller shall always accord most favoured client status to the buyer/consignee and shall give the most competitive price for spares and consumables of its machines/equipments supplied.
  11. **Installation, Training, Manuals:** Seller shall be responsible to carry out Installation & commissioning, Supervision and Demonstration of the goods. They shall provide required jigs and tools for assembly, minor civil works for the completion of the installation and Training of Consignee's representatives for operating and maintaining the equipment and supplying required number of operation & maintenance manual for the goods. In case the category parameters are specifying any requirements regarding the installations, training and manuals the same shall also be applicable.
  12. **Electrical safety checking:** Sellers are required to make sure that they furnish the list of equipments for carrying out routine and preventive maintenance to buyer/consignee .They should make sure to periodically check the electrical safety aspects as per BIS Safety Standards or equivalent .In case they do not have required equipment for such testing should ensure that the

equipments checked for electrical safety compliance through labs with facilities for such checking during every preventive maintenance call.

13. **Software:** All software updates should be provided free of cost during warranty period.

## क्रेता द्वारा जोड़ी गई बिड की विशेष शर्तें/**Buyer Added Bid Specific Terms and Conditions**

1. Experience Certificate for the supply of the same to any Govt/ PSU/ any renowned private organisation along with Supply/ Purchase Order.
2. If the agency is registered under MSME or NSIC, then EMD exemption certificate needs to be enclosed.
3. **Generic**

**Bidder financial standing:** The bidder should not be under liquidation, court receivership or similar proceedings, should not be bankrupt. Bidder to upload undertaking to this effect with bid.

4. **Generic**

Bidders shall quote only those products (Part of Service delivery) in the bid which are not obsolete in the market and has at least 1 years residual market life i.e. the offered product shall not be declared end-of-life by the OEM before this period.

5. **Generic**

Bidders are advised to check applicable GST on their own before quoting. Buyer will not take any responsibility in this regards. GST reimbursement will be as per actuals or as per applicable rates (whichever is lower), subject to the maximum of quoted GST %.

6. **Generic**

Data Sheet of the product(s) offered in the bid, are to be uploaded along with the bid documents. Buyers can match and verify the Data Sheet with the product specifications offered. In case of any unexplained mismatch of technical parameters, the bid is liable for rejection.

7. **Generic**

**End User Certificate:** Wherever Bidders are insisting for End User Certificate from the Buyer, same shall be provided in Buyer's standard format only.

8. **Generic**

Buyer Organization specific Integrity Pact shall have to be complied by all bidders. Bidders shall have to upload scanned copy of signed integrity pact as per Buyer organizations policy along with bid. [Click here to view the file](#)

9. **Generic**

**OPTION CLAUSE:** The Purchaser reserves the right to increase or decrease the quantity to be ordered up to 50 percent of bid quantity at the time of placement of contract. The purchaser also reserves the right to increase the ordered quantity up to 50% of the contracted quantity during the currency of the contract at the contracted rates. The delivery period of quantity shall commence from the last date of original delivery order and in cases where option clause is exercised during the extended delivery period the additional time shall commence from the last date of extended delivery period. The additional delivery time shall be  $(\text{Increased quantity} \div \text{Original quantity}) \times \text{Original delivery period (in days)}$ , subject to minimum of 30 days. If the original delivery period is less than 30 days, the additional time equals the original delivery period. The Purchaser may extend this calculated delivery duration up to the original delivery period while exercising the option clause. Bidders must comply with these terms.

10. **Generic**

**Manufacturer Authorization:** Wherever Authorised Distributors/service providers are submitting the bid, Authorisation Form /Certificate with OEM/Original Service Provider details such as name, designation, address, e-mail Id and Phone No. required to be furnished along with the bid

11. **Generic**

While generating invoice in GeM portal, the seller must upload scanned copy of GST invoice and the screenshot of GST portal confirming payment of GST.

12. **Scope of Supply**

Scope of supply (Bid price to include all cost components) : Supply Installation Testing and Commissioning of Goods

13. **Turnover**

Bidder Turn Over Criteria: The minimum average annual financial turnover of the bidder during the last three years, ending on 31st March of the previous financial year, should be as indicated in the bid document. Documentary evidence in the form of certified Audited Balance Sheets of relevant periods or a certificate from the Chartered Accountant / Cost Accountant indicating the turnover details for the relevant period shall be uploaded with the bid. In case the date of constitution / incorporation of the bidder is less than 3 year old, the average turnover in respect of the completed financial years after the date of constitution shall be taken into account for this criteria.

14. **Turnover**

OEM Turn Over Criteria: The minimum average annual financial turnover of the OEM of the offered product during the last three years, ending on 31st March of the previous financial year, should be as indicated in the bid document. Documentary evidence in the form of certified Audited Balance Sheets of relevant periods or a certificate from the Chartered Accountant / Cost Accountant indicating the turnover details for the relevant period shall be uploaded with the bid. In case the date of constitution / incorporation of the OEM is less than 3 year old, the average turnover in respect of the completed financial years after the date of constitution shall be taken into account for this criteria. In case of bunch bids, the OEM of CATEGORY RELATED TO primary product having highest bid value should meet this criterion.

15. **OEM**

IMPORTED PRODUCTS: In case of imported products, OEM or Authorized Seller of OEM should have a registered office in India to provide after sales service support in India. The certificate to this effect should be submitted.

16. **Service & Support**

Dedicated /toll Free Telephone No. for Service Support : BIDDER/OEM must have Dedicated/toll Free Telephone No. for Service Support.

17. **Service & Support**

Escalation Matrix For Service Support : Bidder/OEM must provide Escalation Matrix of Telephone Numbers for Service Support.

18. **Certificates**

Bidder's offer is liable to be rejected if they don't upload any of the certificates / documents sought in the Bid document, ATC and Corrigendum if any.

19. **Certificates**

The bidder is required to upload, along with the bid, all relevant certificates such as BIS licence, type test certificate, approval certificates and other certificates as prescribed in the Product Specification given in the bid document.

20. **Warranty**

Warranty period of the supplied products shall be 2 years from the date of final acceptance of goods or after completion of installation, commissioning & testing of goods (if included in the scope of supply), at consignee location. OEM Warranty certificates must be submitted by Successful Bidder at the time of

delivery of Goods. The seller should guarantee the rectification of goods in case of any break down during the guarantee period. Seller should have well established Installation, Commissioning, Training, Troubleshooting and Maintenance Service group in INDIA for attending the after sales service. Details of Service Centres near consignee destinations are to be uploaded along with the bid.

#### 21. **Warranty**

Over and above the normal Warranty terms as per GeM GTC, the successful bidder / OEM shall have to provide Comprehensive Warranty during the entire Standard warranty period as per contract. : The comprehensive warranty shall be covering the following scope

CMC cost includes all spares, accessories, all software updation, if required except for single use consumables

(Upload an undertaking with the bid confirming compliance by the bidder if Bidder is taking onus of this compliance. In case OEM is taking onus of this compliance, OEM undertaking is to be uploaded along with Bidder undertaking)

#### 22. **Warranty**

Successful bidder will have to ensure that adequate number of dedicated technical service personals / engineers are designated / deployed for attending to the Service Request in a time bound manner and for ensuring Timely Servicing / rectification of defects during warranty period, as per Service level agreement indicated in the relevant clause of the bid.

#### 23. **Warranty**

Timely Servicing / rectification of defects during warranty period: After having been notified of the defects / service requirement during warranty period, Seller has to complete the required Service / Rectification within 3 days time limit. If the Seller fails to complete service / rectification with defined time limit, a penalty of 0.5% of Unit Price of the product shall be charged as penalty for each week of delay from the seller. Seller can deposit the penalty with the Buyer directly else the Buyer shall have a right to recover all such penalty amount from the Performance Security (PBG). Cumulative Penalty cannot exceed more than 10% of the total contract value after which the Buyer shall have the right to get the service / rectification done from alternate sources at the risk and cost of the Seller besides forfeiture of PBG. Seller shall be liable to re-imburse the cost of such service / rectification to the Buyer.

#### 24. **Past Project Experience**

**Proof for Past Experience and Project Experience clause:** For fulfilling the experience criteria any one of the following documents may be considered as valid proof for meeting the experience criteria: a. Contract copy along with Invoice(s) with self-certification by the bidder that service/supplies against the invoices have been executed. b. Execution certificate by client with contract value. c. Any other document in support of contract execution like Third Party Inspection release note, etc. **Proof for Past Experience and Project Experience clause:** For fulfilling the experience criteria any one of the following documents may be considered as valid proof for meeting the experience criteria: a. Contract copy along with Invoice(s) with self-certification by the bidder that service/supplies against the invoices have been executed. b. Execution certificate by client with contract value. c. Any other document in support of contract execution like Third Party Inspection release note, etc.

#### 25. **Forms of EMD and PBG**

Bidders can also submit the EMD with Payment online through RTGS / internet banking in Beneficiary name

AIIMS GUWAHATI EMD/SD  
Account No.  
497701000072  
IFSC Code  
ICIC0004977  
Bank Name  
ICICI  
Branch address  
CHANGSARI, GUWAHATI  
.

Bidder to indicate bid number and name of bidding entity in the transaction details field at the time of on-line transfer. Bidder has to upload scanned copy / proof of the Online Payment Transfer along with bid.

## 26. **Forms of EMD and PBG**

Successful Bidder can submit the Performance Security in the form of Payment online through RTGS / internet banking also (besides PBG which is allowed as per GeM GTC). On-line payment shall be in Beneficiary name

AIIMS GUWAHATI EMD/SD

Account No.

497701000072

IFSC Code

ICIC0004977

Bank Name

ICICI

Branch address

CHANGSARI, GUWAHATI

. Successful Bidder to indicate Contract number and name of Seller entity in the transaction details field at the time of on-line transfer. Bidder has to upload scanned copy / proof of the Online Payment Transfer in place of PBG within 15 days of award of contract.

## 27. **Financial Criteria**

**NET WORTH:** Net Worth of the OEM should be positive as per the last audited financial statement.

## 28. **Buyer Added Bid Specific ATC**

Buyer Added text based ATC clauses

**1. Technical Data Sheet Submission:** Participant bidders must provide a Technical Data Sheet of the quoted product upon request during technical evaluation. The Evaluation Committee will cross-check this data sheet with the product specifications demonstrated. Any unexplained discrepancies in technical parameters may lead to rejection of the bid of the bidder.

**2. Demonstration Requirement:** The Tender Evaluation Committee (TEC)/ Indenting Department may ask the Technically Qualified Sellers to arrange demonstration of the quoted equipment in physical or virtual mode as decided by the TEC/IO at AIIMS Guwahati site before opening of financial bid/placement of Purchase order. Date and time for demonstration will be communicated as deemed fit by TEC / Indenting department.

**3. Fall Clause Certificate-** Post Technical Evaluation/After opening of Financial Bid, Submission of Fall Clause Certificate by the L- Bidder is mandatory, Non submission of the required certificate may lead to disqualification.

**4. Price Justification/Price Reasonability-** Post Technical Evaluation/After opening of Financial Bid, L-1 Seller has to submit valid documents for Price Justification. Purchase Orders for the items supplied previously to any Govt Organization needs to be submitted.

5. **Integrity Pact-** Only those Vendors/Bidders, who commit themselves to Integrity Pact with the buyer, would be considered competent to participate in the bidding process. Entering into this Pact would be a preliminary qualification. Non submission of the required certificate will lead to disqualification during the Technical Evaluation.

6. **EMD/PBG DETAILS** : ACCOUNT HOLDER NAME - AIIMS GUWAHATI EMD/SD ACCOUNT NO- 497701000072 IFSC CODE- ICIC0004977

Earnest Money Deposit (EMD) must be submitted via [Real Time Gross Settlement \(RTGS\)](#) only. Bidder must use this electronic fund transfer method, rather than cheques or other forms of payment, and then upload proof of the transaction as an attachment being one of the ATC document.

29. **Buyer Added Bid Specific ATC**

Buyer uploaded ATC document [Click here to view the file.](#)

## अस्वीकरण/Disclaimer

The additional terms and conditions have been incorporated by the Buyer after approval of the Competent Authority in Buyer Organization, whereby Buyer organization is solely responsible for the impact of these clauses on the bidding process, its outcome, and consequences thereof including any eccentricity / restriction arising in the bidding process due to these ATCs and due to modification of technical specifications and / or terms and conditions governing the bid. If any clause(s) is / are incorporated by the Buyer regarding following, the bid and resultant contracts shall be treated as null and void and such bids may be cancelled by GeM at any stage of bidding process without any notice:-

1. Definition of Class I and Class II suppliers in the bid not in line with the extant Order / Office Memorandum issued by DPIIT in this regard.
2. Seeking EMD submission from bidder(s), including via Additional Terms & Conditions, in contravention to exemption provided to such sellers under GeM GTC.
3. Publishing Custom / BOQ bids for items for which regular GeM categories are available without any Category item bunched with it.
4. Creating BoQ bid for single item.
5. Mentioning specific Brand or Make or Model or Manufacturer or Dealer name.
6. Mandating submission of documents in physical form as a pre-requisite to qualify bidders.
7. Floating / creation of work contracts as Custom Bids in Services.
8. Seeking sample with bid or approval of samples during bid evaluation process. (However, in bids for [attached categories](#), trials are allowed as per approved procurement policy of the buyer nodal Ministries)
9. Mandating foreign / international certifications even in case of existence of Indian Standards without specifying equivalent Indian Certification / standards.
10. Seeking experience from specific organization / department / institute only or from foreign / export experience.
11. Creating bid for items from irrelevant categories.
12. Incorporating any clause against the MSME policy and Preference to Make in India Policy.
13. Reference of conditions published on any external site or reference to external documents/clauses.
14. Asking for any Tender fee / Bid Participation fee / Auction fee in case of Bids / Forward Auction, as the case may be.
15. Buyer added ATC Clauses which are in contravention of clauses defined by buyer in system generated bid template as indicated above in the Bid Details section, EMD Detail, ePBG Detail and MII and MSE Purchase

Preference sections of the bid, unless otherwise allowed by GeM GTC.

16. In a category based bid, adding additional items, through buyer added additional scope of work/ additional terms and conditions/or any other document. If buyer needs more items along with the main item, the same must be added through bunching category based items or by bunching custom catalogs or bunching a BoQ with the main category based item, the same must not be done through ATC or Scope of Work.

Further, if any seller has any objection/grievance against these additional clauses or otherwise on any aspect of this bid, they can raise their representation against the same by using the Representation window provided in the bid details field in Seller dashboard after logging in as a seller within 4 days of bid publication on GeM. Buyer is duty bound to reply to all such representations and would not be allowed to open bids if he fails to reply to such representations.

## **Additional Clause For Comprehensive Maintenance Charges**

1.CMC shall include preventive maintenance including calibration as per technical/ service /operational manual of the manufacturer, service charges and spares, after satisfactory completion of Warranty. During the CMC period commencing from date of the successful completion of warranty period, Service personnel shall visit each consignee site as recommended in the manufacturer's technical/ service /operational manual, at least once in six months or as per user requirement. Cost of consumables shall not be included in CMC. Further there will be 98% uptime warranty during CMC period on 24 (hrs) X 7 (days) X 365 (days) basis, with penalty, to extend CMC period by double the downtime period.

2.CMC charges to be indicated as percentage of cost of equipment quoted for each year after the warranty period.

3.GST shall be included in the CMC Charges quoted.

4.Cost of CMC will be added for Ranking/Evaluation purpose with depreciation formula. A 10% discounting rate per year shall be applied on CMC Charges for price evaluation on Net Present Value.

5.The payment of CMC will be made on quarterly basis after satisfactory completion of said period, duly certified by end user.

6.While creating a bid or RA, buyers shall indicate whether CMC is required against Yes/No" options. If CMC Charges are included, an option for number of years for CMC required after the warranty period shall be available. Under this option up to 10 years can be chosen for CMC charges beyond warranty period.

7.In case the bid has a provision for CMC, the warranty of the product will also be deemed to have been converted into Comprehensive warranty including preventive maintenance and calibration as per technical/ service /operational manual of the manufacturer, service charges and spares, during the Warranty Period also. Sellers are therefore advised to include the cost of Comprehensive Warranty including spares (excluding consumables) also in product Cost.

8.The CMC functionality shall be available in bid only and no direct RA shall be applicable. In case of bid to R/A decrement rules shall be applicable on total price inclusive of CMC charges. Bunching of products shall not be available while creating bids with CMC charges.

8.1.Buyer shall indicate number of years of warranty by selecting different options available in the field depending on warranty parameter applicable in category parameters for the equipment. No. of years of warranty indicated here shall supersede the warranty period indicated elsewhere in bid or product specifications. The Seller while participating in Bid/RA will get fields to indicate CMC charges as percentage depending on number of years of CMC selected by Buyer. The following shall be applicable, if 5 year CMC selected:

CMC charges for 1st year after warranty period- Percentage to be indicated- A1

CMC charges for 2nd year after warranty period- Percentage to be indicated- A2

CMC charges for 3rd year after warranty period - Percentage to be indicated- A3

CMC charges for 4th year after warranty period - Percentage to be indicated- A4

CMC charges for 5th year after warranty period - Percentage to be indicated- A5

Similarly, A6 to A10 are to be indicated for 6th to 10th year of CMC if applicable.

8.2.The calculation of CMC Charges shall take into account the number of years of warranty and duration of CMC as specified while creating bid.

8.3.In the price evaluation, the system shall provide function to calculate the cost of each equipment by formula indicated below including CMC and then show the inter-se-ranking of the bidders. The following are the variables

(i) Number of years for which CMC required.

(ii) Number of years of product warranty

The formula for calculating total cost including CMC charges shall be as under:

Total Cost for evaluation=

$C+C*\{(A1/100)/(1.10^n)+(A2/100)/(1.10^{n+1})+(A3/100)/(1.10^{n+2})+(A4/100)/(1.10^{n+3})+(A5/100)/(1.10^{n+4})\}$  and so on

C - Cost for equipment quoted and n shall be number of years of product warranty specified.

If 2 year warranty specified, n shall be 2 and if 5 year warranty specified, n shall be 5. A1,A2, A3, A4& A5 shall depend on how many years CMC selected. For 3 year CMC, only A1,A2 and A3 factors are to be taken into account and A4 and A5 will not be applicable.

8.4.CMC charges offered for each subsequent year should be same or higher than preceding year.

8.5.The CMC charges shall be offered within range of 3 to 25% of cost of equipment.

9.Since CMC charges are to be paid only later for each year during CMC period,applicable performance guarantee amount after placement of contract shall be based on the cost of equipment excluding the cost of CMC Charges.

10.Performance bank guarantee applicable for CMC is to be submitted at start of the CMC and shall be applicable between 2.5% to 10% as specified in bid on total CMC Charges.The PBG submitted after award of contract shall be released only after new PBG for the CMC period is submitted and accepted by buyer/consignee after due verification.Bank guarantee for CMC is to remain valid till completion of CMC period plus one year. The bank guarantee for CMC shall be submitted to buyer directly. In case, seller fails to submit the PBG or does not provide services for the CMC contract after expiry of warranty period then PBG of equipment shall be forfeited.

11.In case of splitting of order quantity, equipment cost and CMC charges offered by L1 bidder shall be matched by higher quoting eligible bidders on one-to-one basis.The equipment cost and CMC charges (year to year) shall be matched individually.

12.The CMC Contract shall be an offline contract to be handled by buyer.The payment of CMC will be made on quarterly basis after satisfactory completion of said period, duly certified by end user and scope of CMC will be as per para 1 above.

13.CMC Charges are inclusive of all the charges for Transportation, Lodging, Boarding, all insurances including third party insurance and all other incidental charges. The same shall include GST. The prices also include cost of spares and damaged parts. Purchaser does not have any liability, whatsoever, over and above the cost of CMC. It also includes for arranging hand tools & tackles, special tools etc. required to carry out the work.

**All GeM Sellers/Service Providers shall ensure full compliance with all applicable labour laws, including the provisions, rules, schemes and guidelines under the four Labour Codes i.e. the Code on Wages, 2019; the Industrial Relations Code, 2020; the Occupational Safety, Health and Working Conditions Code, 2020; and the Code on Social Security, 2020 as and when notified and brought into force by the Government of India.**

**For all provisions of the Labour Codes that are pending operationalisation through rules, schemes or notifications, the corresponding provisions of the pre-existing labour enactments (such as The Minimum Wages Act, 1948, The Payment of Wages Act, 1936, The Payment of Bonus Act, 1965, The Equal Remuneration Act, 1976, The Payment of Gratuity Act, 1972, etc. and relevant State Rules) shall continue to remain applicable.**

**The Seller/ Service Providers shall, therefore, be responsible for ensuring compliance under:**

- **All notified and enforceable provisions of the new Labour Codes as mentioned hereinabove; and**
- **All operative provisions of the erstwhile Labour Laws until their complete substitution.**

**All obligations relating to wages, social security, safety, working conditions, industrial relations etc. and any other statutory requirements shall be strictly met by the Seller/ Service Provider. Any non-compliance shall constitute a breach of the contract and shall entitle the Buyer to take appropriate action in accordance with the contract and applicable law.**

[यह बिड सामान्य शर्तों के अंतर्गत भी शासित है /This Bid is also governed by the General Terms and Conditions](#)

जेम की सामान्य शर्तों के खंड 26 के संदर्भ में भारत के साथ भूमि सीमा साझा करने वाले देश के बिडर से खरीद पर प्रतिबंध के संबंध में भारत के साथ भूमि सीमा साझा करने वाले देश का कोई भी बिडर इस निविदा में बिड देने के लिए तभी पात्र होगा जब वह बिड देने वाला सक्षम प्राधिकारी के पास पंजीकृत हो।बिड में भाग लेते समय बिडर को इसका अनुपालन करना होगा और कोई भी गलत घोषणा किए जाने व इसका अनुपालन न करने पर अनुबंध को तत्काल समाप्त करने और कानून के अनुसार आगे की कानूनी कार्यवाई का आधार होगा।/In terms

of GeM GTC clause 26 regarding Restrictions on procurement from a bidder of a country which shares a land border with India, any bidder from a country which shares a land border with India will be eligible to bid in this tender only if the bidder is registered with the Competent Authority. While participating in bid, Bidder has to undertake compliance of this and any false declaration and non-compliance of this would be a ground for immediate termination of the contract and further legal action in accordance with the laws.

---धन्यवाद/Thank You---