

बिड दस्तावेज़ / Bid Document

बिड विवरण/Bid Details	
बिड बंद होने की तारीख/समय /Bid End Date/Time	05-06-2026 16:00:00
बिड खुलने की तारीख/समय /Bid Opening Date/Time	05-06-2026 16:30:00
बिड पेशकश वैधता (बंद होने की तारीख से)/Bid Offer Validity (From End Date)	180 (Days)
मंत्रालय/राज्य का नाम/Ministry/State Name	Pmo
विभाग का नाम/Department Name	Department Of Atomic Energy
संगठन का नाम/Organisation Name	N/a
कार्यालय का नाम/Office Name	Tata Memorial Centre - Advanced Centre For Treatment, Research And Education In Cancer (actrec)
कुल मात्रा/Total Quantity	1
वस्तु श्रेणी /Item Category	Low Temperature Hydrogen Peroxide Gas Plasma Sterilizer (Q2)
एमएसएमई के लिए अनुभव के वर्षों और टर्नओवर से छूट प्रदान की गई है/MSE Relaxation for Years of Experience and Turnover	No
स्टार्टअप के लिए अनुभव के वर्षों और टर्नओवर से छूट प्रदान की गई है /Startup Relaxation for Years of Experience and Turnover	No
विक्रेता से मांगे गए दस्तावेज़/Document required from seller	Certificate (Requested in ATC),OEM Authorization Certificate,Compliance of BoQ specification and supporting document *In case any bidder is seeking exemption from Experience / Turnover Criteria, the supporting documents to prove his eligibility for exemption must be uploaded for evaluation by the buyer
क्या आप निविदाकारों द्वारा अपलोड किए गए दस्तावेजों को निविदा में भाग लेने वाले सभी निविदाकारों को दिखाना चाहते हैं? संदर्भ मेन् है/Do you want to show documents uploaded by bidders to all bidders participated in bid?	Yes (Documents submitted as part of a clarification or representation during the tender/bid process will also be displayed to other participated bidders after log in)
बिड लगाने की समय सीमा स्वतः नहीं बढ़ाने के लिए आवश्यक बिड की संख्या। / Minimum number of bids required to disable automatic bid extension	3

बिड विवरण/Bid Details	
दिनों की संख्या, जिनके लिए बिड लगाने की समय-सीमा बढ़ाई जाएगी। / Number of days for which Bid would be auto-extended	7
ऑटो एक्सटेंशन अधिकतम कितनी बार किया जाना है। / Number of Auto Extension count	1
बिड से रिवर्स नीलामी सक्रिय किया/Bid to RA enabled	Yes
रिवर्स नीलामी योग्यता नियम/RA Qualification Rule	H1-Highest Priced Bid Elimination
बिड का प्रकार/Type of Bid	Two Packet Bid
तकनीकी मूल्यांकन के दौरान तकनीकी स्पष्टीकरण हेतु अनुमत समय /Time allowed for Technical Clarifications during technical evaluation	3 Days
निरीक्षण आवश्यक (सूचीबद्ध निरीक्षण प्राधिकरण /जेम के साथ पूर्व पंजीकृत एजेंसियों द्वारा)/Inspection Required (By Empanelled Inspection Authority / Agencies pre-registered with GeM)	No
मूल्यांकन पद्धति/Evaluation Method	Total value wise evaluation
मध्यस्थता खंड/Arbitration Clause	No
सुलह खंड/Mediation Clause	No

ईएमडी विवरण/EMD Detail

एडवाइजरी बैंक/Advisory Bank	State Bank of India
ईएमडी राशि/EMD Amount	50000

ईपीबीजी विवरण /ePBG Detail

एडवाइजरी बैंक/Advisory Bank	State Bank of India
ईपीबीजी प्रतिशत (%) /ePBG Percentage(%)	5.00
ईपीबीजी की आवश्यक अवधि (माह) /Duration of ePBG required (Months).	64

(a). जेम की शर्तों के अनुसार ईएमडी छूट के इच्छुक बिडर को संबंधित कटेगरी के लिए बिड के साथ वैध समर्थित दस्तावेज प्रस्तुत करने है। एमएसई कटेगरी के अंतर्गत केवल वस्तुओं के लिए विनिर्माता तथा सेवाओं के लिए सेवा प्रदाता ईएमडी से छूट के पात्र हैं। व्यापारियों को इस नीति के दायरे से बाहर रखा गया है।/EMD EXEMPTION: The bidder seeking EMD exemption, must submit the valid supporting document for the relevant category as per GeM GTC with the bid. Under MSE category, only manufacturers for goods and Service Providers for Services are eligible for exemption from EMD. Traders are excluded from the purview of this Policy.

(b). ईएमडी और संपादन जमानत राशि, जहां यह लागू होती है, लाभार्थी के पक्ष में होनी चाहिए। / EMD & Performance security should be in favour of Beneficiary, wherever it is applicable.

लाभार्थी /Beneficiary :

Director Actrec
Tata Memorial Centre - Advanced Centre For Treatment, Research And Education In Cancer (actrec), Department of Atomic Energy, N/A, PMO
(Director, Actrec)

बोली विभाजन लागू नहीं किया गया/ Bid splitting not applied.

एमआईआई खरीद वरीयता / MII Purchase Preference

एमआईआई खरीद वरीयता / MII Purchase Preference	Yes
मेक इन इंडिया विक्रेताओं को खरीद में प्राथमिकता, यदि उनका मूल्य L1+X% तक की सीमा में है / Purchase Preference to MII sellers available upto price within L1+X%	20
मेक इन इंडिया खरीद में प्राथमिकता के लिए बिड की मात्रा का अधिकतम प्रतिशत / Maximum Percentage of Bid quantity for MII purchase preference	50
सार्वजनिक खरीद (मेक-इन-इंडिया को प्राथमिकता) आदेश 2017 के अनुसार केवल क्लास 1/क्लास 2 के स्थानीय आपूर्तिकर्ताओं को ही भागीदारी की अनुमति है दिनांक 16.09.2020 (समय-समय पर संशोधित एवं लागू) / Allow participation only from Class 1/Class 2 local suppliers as per the Public procurement(Preference to Make-in-india) order 2017 date 16.09.2020(as amended and applicable time to time)	Yes, in compliance with the MII ORDER : DPIIT Order(as amended and applicable time to time)

एमएसई खरीद वरीयता/MSE Purchase Preference

एमएसई खरीद वरीयता/MSE Purchase Preference	Yes
सूक्ष्म और लघु उद्यम मूल उपकरण निर्माताओं को खरीद में प्राथमिकता, यदि उनका मूल्य L1+X% तक की सीमा में हो / Purchase Preference to MSE OEMs available upto price within L1+X%	15
सूक्ष्म और लघु उद्यम को खरीद में प्राथमिकता के लिए बिड की मात्रा का अधिकतम प्रतिशत / Maximum Percentage of Bid quantity for MSE purchase preference	25

1. Preference to Make In India products (For bids < 200 Crore):Preference shall be given to Class 1 local supplier as defined in public procurement (Preference to Make in India), Order 2017 as amended from time to time and its subsequent Orders/Notifications issued by concerned Nodal Ministry for specific Goods/Products. The minimum local content to qualify as a Class 1 local supplier is denoted in the bid document. If the bidder wants to avail the Purchase preference, the bidder must upload a certificate from the OEM regarding the percentage of the local content and the details of locations at which the local value addition is made along with their bid, failing which no purchase preference shall be granted. In case the bid value is more than Rs 10 Crore, the declaration relating to percentage of local content shall be certified by the statutory auditor or cost auditor, if the OEM is a company and by a practicing cost accountant or a chartered accountant for OEMs other than companies as per the Public

Procurement (preference to Make-in -India) order 2017 dated 04.06.2020. Only Class-I and Class-II Local suppliers as per MII order dated 4.6.2020 will be eligible to bid. Non - Local suppliers as per MII order dated 04.06.2020 are not eligible to participate. However, eligible micro and small enterprises will be allowed to participate .The buyers are advised to refer the OM No.F.1/4/2021-PPD dated 18.05.2023.

[OM No.1 4 2021 PPD dated 18.05.2023](#) for compliance of Concurrent application of Public Procurement Policy for Micro and Small Enterprises Order, 2012 and Public Procurement (Preference to Make in India) Order, 2017.

2. Purchase preference will be given to MSEs having valid Udyam Registration and whose credentials are validated online through Udyam Registration portal as defined in Public Procurement Policy for Micro and Small Enterprises (MSEs) Order, 2012 dated 23.03.2012 issued by Ministry of Micro, Small and Medium Enterprises and its subsequent Orders/Notifications issued by concerned Ministry. If the bidder wants to avail themselves of the Purchase preference, the bidder must be the manufacturer / OEM of the offered product on GeM. Traders are excluded from the purview of Public Procurement Policy for Micro and Small Enterprises and hence resellers offering products manufactured by some other OEM are not eligible for any purchase preference. In respect of bid for Services, the bidder must be the Service provider of the offered Service. Relevant documentary evidence in this regard shall be uploaded along with the bid in respect of the offered product or service and Buyer will decide eligibility for purchase preference based on documentary evidence submitted, while evaluating the bid. If L-1 is not an MSE and MSE Seller (s) has / have quoted price within L-1+ 15% (Selected by Buyer) of margin of purchase preference /price band defined in relevant policy, such MSE Seller shall be given opportunity to match L-1 price and contract will be awarded for 25% (selected by Buyer) percentage of total quantity. The buyers are advised to refer the OM No. F.1/4/2021-PPD dated 18.05.2023 [OM No.1 4 2021 PPD dated 18.05.2023](#) for compliance of Concurrent application of Public Procurement Policy for Micro and Small Enterprises Order, 2012 and Public Procurement (Preference to Make in India) Order, 2017. Benefits of MSE will be allowed only if seller is validated on-line in GeM profile as well as validated and approved by Buyer after evaluation of documents submitted.

3. Estimated Bid Value indicated above is being declared solely for the purpose of guidance on EMD amount and for determining the Eligibility Criteria related to Turn Over, Past Performance and Project / Past Experience etc. This has no relevance or bearing on the price to be quoted by the bidders and is also not going to have any impact on bid participation. Also this is not going to be used as a criteria in determining reasonableness of quoted prices which would be determined by the buyer based on its own assessment of reasonableness and based on competitive prices received in Bid / RA process.

4. Reverse Auction would be conducted amongst all the technically qualified bidders except the Highest quoting bidder. The technically qualified Highest Quoting bidder will not be allowed to participate in RA. However, H-1 will also be allowed to participate in RA in following cases:

- i. If number of technically qualified bidders are only 2 or 3.
- ii. If Buyer has chosen to split the bid amongst N sellers, and H1 bid is coming within N.
- iii. In case Primary product of only one OEM is left in contention for participation in RA on elimination of H-1.
- iv. If L-1 is non-MSE and H-1 is eligible MSE and H-1 price is coming within price band of 15% of Non-MSE L-1
- v. If L-1 is non-MII and H-1 is eligible MII and H-1 price is coming within price band of 20% of Non-MII L-1

Low Temperature Hydrogen Peroxide Gas Plasma Sterilizer (1 pieces)

(क्रमशः श्रेणी 1 और श्रेणी 2 के स्थानीय आपूर्तिकर्ता के रूप में अर्हता प्राप्त करने के लिए आवश्यक/Minimum 50% and 20% Local Content required for qualifying as Class 1 and Class 2 Local Supplier respectively)

तकनीकी विशिष्टियाँ /Technical Specifications

[* जेम कटेगरी विशिष्टि के अनुसार / As per GeM Category Specification](#)

विवरण/Specification	विशिष्टि का नाम /Specification Name	बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)
GENERAL	Product Description	Low temperature Hydrogen Peroxide Gas Plasma Sterilizer
PRODUCT INFORMATION	Type of Hydrogen Peroxide Gas Sterilisation Technology	Plasma technology
	Sterilizer is environment friendly and have no toxic by products or harmful residuals	Yes

विवरण/Specification	विशिष्टि का नाम /Specification Name	बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)
	Shape of Sterilization Chamber	Rectangular
	Usable capacity of Sterilization Chamber	76 to 100 liters
	Sterilization Chamber Door	Pneumatically or electrically controlled door with complete safety protocol with multiple locking.
	Control System	Microprocessor controlled system with clear user interface for control and display of cycle phases, parameters values and text messages.
	Sterilizer has facility to completely monitor its operations with audiovisual alarms and alarm history	Yes
	Built in facility for recording and printing cycle details	Yes
	Mobility	Mobile on castor wheel and compact in design
	Cycle temperature range	35 degree to 60 degree centigrade
	Have pre programmed sterilisation cycles for flat instruments, and for hollow and flexible items	Yes
	Built in safety features available	Emergency stop, Door safety, Over heat protection, Electrical safety
	Have an easy to troubleshoot by display interface and all services access available from the front	Yes
	Interface for remote supervision and data storage	Digital interface for remote supervision and data storage
	Consumables be provided with each Sterilisation Unit sufficient for running full load Sterilization cycles(at least)	200
	Sterilizing Chamber has sufficient dimensions for sterilizing rigid telescope units	Yes
	Sterilization Cycle time (± 5 Minutes)	60 Minutes
	Capable to distribute the plasma homogenously through the Chamber	Yes

विवरण/Specification	विशिष्टि का नाम /Specification Name	बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)
	Have diagnostic system to detects & remove excess moisture thus eliminating chances of contamination due to residual bio burden prior to sterilization cycles & self-test functions	Yes
	Type of Display	LCD/LED/TFT
	Power Supply	220-240V/50Hz
CERTIFICATIONS	Compliance to Medical Device Rules (MDR) 2017 as amended till date	Yes
	Availability of valid Medical Device license for the product issued from the competent authority defined under Drugs and Cosmetic Act 1940 and Rules made there under as amended till date	Yes
	Availability of Test Report for each supplied batch/product as per Medical Device Rule (MDR) 2017 as amended till date	Yes
	Submission of all necessary certifications, licenses and test reports to the buyer at the time of bid submission or along with supplies as per buyer requirement	Yes
WARRANTY	Warranty in years (Option of comprehensive warranty is available through bidding only, which if opted will supersede normal warranty in the catalogue)	5 Or higher (year)

Additional Specification Parameters - Low Temperature Hydrogen Peroxide Gas Plasma Sterilizer (1 pieces)

Specification Parameter Name	Bid Requirement (Allowed Values)
1. Sterilization Process 2. Compatibility 3. Sterilant 4. Sterilization Monitoring & Alarms 5. Documentation	1. Simple & fast low temperature sterilization with the help of H2O2 & effective removal of H2O2 from sterilized items & to compliment the process 2 Suitable for heat-sensitive instruments, including rigid endoscopes and lumens. 3. should be in a cassette/ bottle /cartridge/equivalent with H2O2 concentration $\geq 55\%$ Should be able to store at room temperature with leak indicator. 4. Full operation monitoring with audiovisual alarms and cycle history. 5. Integrated data recording and printer required.

Specification Parameter Name	Bid Requirement (Allowed Values)
1.MDM Endorsements 2. Conditioning 3. The system should use minimum quantity of sterilant	1.Must be recommended by at least 3 major medical device manufacturers (e.g., Storz, Olympus for sterilization of Telescopes and camera heads. Document is required from medical device Manufacturer only 2. Moisture detection phase/ load conditioning phase should be available in the equipment prior to sterilant injection phase to reduce aborts due to moisture and avoids sterilant wastage. Valid documents to be provided. 3.Uses < 8 ml of sterilant per injection for dry, non-corrosive cycles.
1.Sterilizer Ergonomics 2. Shelf Life of sterilant 3. Consumables (Sterilant) to be quoted in the financial bid 4. Provide the list of proprietary consumables if any for the quoted equipment in technical bid (Do not mention Price in technical bid)	1.Should Have Ergonomic foot pedal or door sensor for hands-free operation.2. Sterilant must have a 10-month shelf life upon delivery. 3. Quote consumables based on an estimated 1,800 annual cycles. (Number of Sterilization cycles for 1 year is an estimate actual Number cycles may vary) 4. Provide the list of proprietary consumables if any for the quoted equipment along with the technical bid. If no list is provided, it implies that there are no proprietary consumables.
1.Experience 2.Biological indicator (BI): Reader with readout time of minimum 30 minutes. Datasheet for BI reader to be uploaded along with the bid. Compliance sheet needs to be uploaded for same. 3. Accessories	1. Vendor should have experience in selling Plasma/H2O2 Sterilizer equipment in the country for last 5 years. Quoted model also should have installations in India. 2. Please Specify & Please Submit Compliance sheet with all the relevant documents 3. Sterilizer should be provided with suitable power rating servo sterilizer and biological incubator (BI) and Reader with readout time of minimum 30 minutes.
1.Certifications 2. ENISO 14937- Process certificate 3. Lumen sterilization Validation	1. quoted model should be indicated for sterilization of metal and non-metal medical devices by USFDA (510K)or European CE or ICMED or BIS 2. ENISO 14937- Process certificate should be available 3. Lumen sterilization claims should be validated & endorsed by US FDA (510K) or European CE or ICMED or BIS
1.Sterilization Process 2. Compatibility 3. Sterilant 4. Sterilization Monitoring & Alarms 5. Documentation	1.Simple & fast low temperature sterilization with the help of H2O2 & effective removal of H2O2 from sterilized items & to compliment the process 2 Suitable for heat-sensitive instruments, including rigid endoscopes and lumens. 3.hould be in a cassette/ bottle /cartridge/equivalent with H2O2 concentration>= 55% Should be able to store at room temperature with leak indicator. 4. Full operation monitoring with audiovisual alarms and cycle history. 5.Integrated data recording and printer required.
1.MDM Endorsements 2. Conditioning 3. The system should use minimum quantity of sterilant	1.Must be recommended by at least 3 major medical device manufacturers (e.g., Storz, Olympus for sterilization of Telescopes and camera heads. Document is required from medical device Manufacturer only 2. Moisture detection phase/ load conditioning phase should be available in the equipment prior to sterilant injection phase to reduce aborts due to moisture and avoids sterilant wastage. Valid documents to be provided. 3.Uses < 8 ml of sterilant per injection for dry, non-corrosive cycles.
1.Sterilizer Ergonomics 2. Shelf Life of sterilant 3. Consumables (Sterilant) to be quoted in the financial bid 4. Provide the list of proprietary consumables if any for the quoted equipment in technical bid (Do not mention Price in technical bid)	1.Should Have Ergonomic foot pedal or door sensor for hands-free operation.2. Sterilant must have a 10-month shelf life upon delivery. 3. Quote consumables based on an estimated 1,800 annual cycles. (Number of Sterilization cycles for 1 year is an estimate actual Number cycles may vary) 4. Provide the list of proprietary consumables if any for the quoted equipment along with the technical bid. If no list is provided, it implies that there are no proprietary consumables.

Specification Parameter Name	Bid Requirement (Allowed Values)
1.Experience 2.Biological indicator (BI): Reader with readout time of minimum 30 minutes. Datasheet for BI reader to be uploaded along with the bid. Compliance sheet needs to be uploaded for same. 3. Accessories	1. Vendor should have experience in selling Plasma/H2O2 Sterilizer equipment in the country for last 5 years. Quoted model also should have installations in India. 2. Please Specify & Please Submit Compliance sheet with all the relevant documents 3. Sterilizer should be provided with suitable power rating servo sterilizer and biological incubator (BI) and Reader with readout time of minimum 30 minutes.
1.Certifications 2. ENISO 14937- Process certificate 3. Lumen sterilization Validation	1. quoted model should be indicated for sterilization of metal and non-metal medical devices by USFDA (510K)or European CE or ICMED or BIS 2. ENISO 14937- Process certificate should be available 3. Lumen sterilization claims should be validated & endorsed by US FDA (510K) or European CE or ICMED or BIS

* Bidders offering must also comply with the additional specification parameters mentioned above.

परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.No.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Kanchana Gopalkrishnan	410210,Tata Memorial Center - Advanced Centre for Treatment, Research and Education in Cancer (ACTREC), Plt no.1 &2, Sector 22, Kharghar, Navi Mumbai	1	45

Special terms and conditions-Version:1 effective from 01-08-2024 for category Low Temperature Hydrogen Peroxide Gas Plasma Sterilizer

1. All Provisions of Drugs and Cosmetics Act, 1940 and Rules (including Medical Device Rule 2017) made there under as amended till date will always be applicable. This will include all notifications issued by Central Drugs Standard Control Organization (CDSCO), Ministry of Health & Family Welfare (MoHFW) and Department of Pharmaceuticals (DOP), Ministry of Chemicals & Fertilizers time to time in this regard.
2. The sellers are registered on GeM based on self declaration of valid Medical Device License, product certification, test reports etc. However, buyers must check and validate the details at their end for all applicable licenses and certifications e.g., validity and authenticity/genuineness of Medical Device license, product certification, manufacturer certification/licenses, test reports etc.
3. In case of authorized resellers/distributors, it will be the legal & regulatory liability of the manufacturer to ensure that their resellers/distributors are operating in compliance with all relevant laws and regulations and are properly licensed to sell the manufacturer's products, including verifying the validity and authenticity of Medical Device license held by them.
4. The price offered by the seller/bidder shall not, in any case exceed the DPCO/NPPA controlled price or price fixed by State Government, if any. The seller must reduce the prices if there is any reduction in DPCO/NPPA ceiling price or price fixed by State Government, if any.
5. Any other Terms and Conditions which is not included or at variance with the conditions specified in STC/GTC, may be added by the buyer through Additional Terms and Conditions (ATC) in the bid to ensure items are procured from authentic/validated source with appropriate and applicable quality. The above terms and conditions are in reverse order of precedence i.e. ATC shall supersede specific STC which shall supersede General Terms and Conditions (GTC), whenever there are any conflicting provisions.
6. **Comprehensive warranty:** Comprehensive warranty shall include preventive maintenance including calibration as per technical/ service /operational manual of the manufacturer, service charges and spares. During the warranty period commencing from date of the successful

completion of warranty period, Service personnel shall visit each consignee site as recommended in the manufacturer's technical/ service /operational manual, at least once in six months. warranty shall not be including the consumables. Further there will be 98% uptime warranty during warranty period on 24 (hrs) X 7 (days) X 365 (days) basis, with penalty, to extend warranty period by double the downtime period.

7. **Service centres:** Details of Service outlets in India to render services for equipment to be furnished to buyer/consignees with complete address, telephone numbers, e mails etc at time of making the supplies. It shall be the responsibility of seller to ensure that authorized service centres are available to cater to the areas where supplies are made within reasonable distance from where the service calls can be handled. Details of toll-free numbers for service call and online registration of service requests also to be provided buyer/consignee at the time of supplies.
8. **Source of supply:** It shall be responsibility of seller to provide Documents regarding source of equipments such as copy of Performa invoice or any other documents to establish that the products supplied are manufactured by OEM indicated and sourced from them.
9. **Packing and Marking:** Medical equipments being very delicate and sensitive packing for the goods should be strong and durable enough to withstand transit including transshipment (if any), rough handling, open storage etc. without any damage, deterioration etc. .The size, weights and volumes of the packing cases, remoteness of the final destination of the goods, availability or otherwise of transport and handling facilities at all points during transit up to final destination,. Quality of packing, the manner of marking within & outside the packages and provision of accompanying documentation shall take into consideration the type of medical equipments being supplied. The accessories shall be suitably labelled and packed. Each of the package shall be marked on three sides with indelible paint of proper quality: indicating contract number and date, brief description of goods including quantity, Packing list reference number, country of origin of goods and any other relevant details.
10. **Spare Parts:** Seller shall provide materials, information etc. pertaining to spare parts manufactured and supplied by the OEM. It shall be ensured that the required spares are available for purchase at least for 10 years from date of supplies. In case due to any reasons the production of the spare parts is discontinued sufficient advance notice should be given to the buyer/consignee before such discontinuation to provide adequate time to purchase the required spare parts etc. Further, OEM and their service centres/dealers shall carry sufficient inventories to assure ex-stock supply of consumables and spares for the equipments so that the same are available. OEM or reseller shall always accord most favoured client status to the buyer/consignee and shall give the most competitive price for spares and consumables of its machines/equipments supplied.
11. **Installation, Training, Manuals:** Seller shall be responsible to carry out Installation & commissioning, Supervision and Demonstration of the goods. They shall provide required jigs and tools for assembly, minor civil works for the completion of the installation and Training of Consignee's representatives for operating and maintaining the equipment and supplying required number of operation & maintenance manual for the goods. In case the category parameters are specifying any requirements regarding the installations, training and manuals the same shall also be applicable.
12. **Electrical safety checking:** Sellers are required to make sure that they furnish the list of equipments for carrying out routine and preventive maintenance to buyer/consignee .They should make sure to periodically check the electrical safety aspects as per BIS Safety Standards or equivalent .In case they do not have required equipment for such testing should ensure that the equipments checked for electrical safety compliance through labs with facilities for such checking during every preventive maintenance call.
13. **Software:** All software updates should be provided free of cost during warranty period.

क्रेता द्वारा जोड़ी गई बिड की विशेष शर्तें/**Buyer Added Bid Specific Terms and Conditions**

1. **Service & Support**

Availability of Service Centres: Bidder/OEM must have a Functional Service Centre in the State of each Consignee's Location in case of carry-in warranty. (Not applicable in case of goods having on-site warranty). If service center is not already there at the time of bidding, successful bidder / OEM shall have to establish one within 30 days of award of contract. Payment shall be released only after submission of documentary evidence of having Functional Service Centre.

2. **Generic**

Bidder shall submit the following documents along with their bid for Vendor Code Creation:

- a. Copy of PAN Card.
- b. Copy of GSTIN.
- c. Copy of Cancelled Cheque.
- d. Copy of EFT Mandate duly certified by Bank.

3. Buyer Added Bid Specific ATC

Buyer Added text based ATC clauses

1. After opening of the Technical bid (Part-1), Physical demonstration of the quoted model may have to be shown / arranged by the bidder, if requested by the Institute. Physical demonstration may be shown at one of the end user’s site/Principle company’s application lab/ manufacturing site located in MMR Region. If there are no installations of the quoted model in MMR Region, then the quoted model may have to be brought in at ACTREC for demo purpose within 10 days from the date of request. Physical Demonstration may be requested to confirm the availability of any or all technical features as mentioned/stated in the technical bid. Physical Demonstration will also be a part of technical evaluation process. If the bidder does not comply, such bids are liable to be disqualified. (Demonstration of quoted model is to be shown and not the demonstration of similar models with different technical specifications and features)

2. Submission of Main Equipment Rate and Unit Rates of Consumable valid for the period of 2 Years:

a. Bidder MUST upload the filled format as detailed in the below table as a supporting PDF file along with the financial bid on GeM. Bidder has to fill the total figure as reflected in last Row of below table in the financial quote field on GeM. Non submission of supporting PDF in exact same format as below shall result in rejection of the financial bid of the bidder and Tata Memorial Centre-ACTREC reserves the right to treat the next successful bidder as L1 for a ward of contract through GeM or outside GeM

Sr. No	Item Description	Quantity (A)	UOM	Unit Rate inclusive of GST (in Rs.) (B)	Total Value in Rs. (C = A X B)
1	Plasma Sterilizer	01	Nos.		
2	Sterilant	01	Nos.		
Total Value of all above items in Rs.					

b. Rates of above consumables shall be valid for 2 Years period from the date of issue of Or

der for Main Equipment, without any escalation and Tata Memorial Centre-ACTREC reserves the right to issue separate orders on as and when required basis for these consumables at the rates mentioned in uploaded PDF file along with the financial bid on GeM

c. Anticipated Quarterly Consumption of above mentioned individual consumable is mentioned in Technical Specification, however Tata Memorial Centre-ACTREC does not guarantee issuance of any Order/s and same shall be on as and when required basis

d. While submitting financial bid on GeM for Main equipment, bidder has to quote considering quantity of individual consumable per quarter as mentioned in Technical Specification and in the event of order, these quantities shall be supplied along with Equipment

3. It is mandatory to upload Technical Catalogue, Brochure, Literature, Specification sheet etc . of the quoted item along with bid
4. It is mandatory to upload actual product image or drawing of quoted item, clearly depicting internal and external dimensions

5 Submission of Price Elements: Bidders must submit all price-related information exclusively in the *Financial Bid*. Any mention of price elements in the *Technical Bid* will render the offer non-responsive and liable for rejection, except in the case of **Single Packet Bidding**.

6. Eligibility Criteria for Bidders: The bidder must be one of the following:

- The manufacturer,
- A 100% subsidiary of a foreign manufacturer operating in India,
- The sole authorized distributor of the manufacturer, or
- A direct authorized distributor of the manufacturer.

Note: Sub-authorization (i.e., authorization granted by an intermediary rather than the manufacturer) will not be accepted.

Submission of Manufacturer Authorization: Where authorized distributors submit the bid, a bid-specific, valid Authorization Form/Certificate from the OEM must be furnished along with the bid. This document should include OEM details such as name, designation, address, email ID, and phone number. Failure to provide this will result in summary rejection of the offer.

7. Physical delivery and installation shall be at ICTREC , Tata Memorial Centre, Survey no. 1A/1/A, Tambati village, Khalapur Taluka, Raigad district, Maharashtra 410203 . The quoted price must include all transportation and incidental charges to this site, regardless of the consignee address shown on the GeM bid.

4. **OEM**

IMPORTED PRODUCTS: In case of imported products, OEM or Authorized Seller of OEM should have a registered office in India to provide after sales service support in India. The certificate to this effect should be submitted.

5. **Scope of Supply**

Scope of supply (Bid price to include all cost components) : Supply Installation Testing Commissioning of Goods and Training of operators and providing Statutory Clearances required (if any)

6. **Warranty**

Warranty period of the supplied products shall be 5 years from the date of final acceptance of goods or after completion of installation, commissioning & testing of goods (if included in the scope of supply), at consignee location. OEM Warranty certificates must be submitted by Successful Bidder at the time of delivery of Goods. The seller should guarantee the rectification of goods in case of any break down during the guarantee period. Seller should have well established Installation, Commissioning, Training, Troubleshooting and Maintenance Service group in INDIA for attending the after sales service. Details of Service Centres near consignee destinations are to be uploaded along with the bid.

7. **Generic**

Installation, Commissioning, Testing, Configuration, Training (if any - which ever is applicable as per scope of supply) is to be carried out by OEM / OEM Certified resource or OEM authorised Reseller.

8. **Generic**

Manufacturer Authorization: Wherever Authorised Distributors/service providers are submitting the bid, Authorisation Form /Certificate with OEM/Original Service Provider details such as name, designation, address, e-mail Id and Phone No. required to be furnished along with the bid

9. **Forms of EMD and PBG**

Bidders can also submit the EMD with Account Payee Demand Draft in favour of

Director actrec
payable at
Tata memorial Centre ACTREC Kharghar Navi Mumbai 410210

Bidder has to upload scanned copy / proof of the DD along with bid and has to ensure delivery of hardcopy to the Buyer within 5 days of Bid End date / Bid Opening date.

अस्वीकरण/Disclaimer

The Additional Terms and Conditions (ATC) have been incorporated by the Buyer after approval of their Competent Authority. The Buyer ,is solely responsible for the impact of these clauses on the bidding process, its outcome, and consequences thereof including any restriction arising in the bidding process due to these ATCs and including the modification of technical specifications and / or terms and conditions governing the bid.All representations / grievances pertaining to the ATC clauses shall be raised with the buyer organization directly and not with GeM.If any of the clause(s) is/are incorporated by the Buyer regarding the following, the bid & resultant contract shall be treated as null & void. Further, GeM reserves the right, at its sole discretion, to cancel the bid forthwith, without issuance of any prior notice or intimation :-

1. Publishing Custom / BOQ bids for items for which regular GeM categories are available (unless such Custom / BOQ item is bunched with the major regular product Category Item).
2. Mandating procurement of / from specific Brand / Make / Model / Manufacturer / Dealer except in case of Single Bid / Proprietary Article Certificate (PAC) Buying.
3. Inclusion of disqualification criteria related to suspension of seller / service provider, where such suspension period has already expired.
4. Mandating submission of documents in physical form as a pre-requisite to qualify bidders.
5. Publishing bids on GeM for procurement of works.
6. Procurement of Goods by creating a Service bid on GeM & vice-versa.
7. Seeking sample with bid or approval of samples during bid evaluation process. However, trial / sample, as the case may be, shall be permitted in cases where trial / sample are allowed as per approved and published procurement policy of the Buyers' controlling Ministry / Department / State / Public Sector Enterprises Headquarters. If there is any violation of trial / sample clause with regard to approved policy of the Buyers' Ministry / Department / State / Public Sector Enterprises Headquarters, then this is to be determined and redressed by the concerned Buyer Organisation only.
8. Seeking experience from specific organization / department / institute only or from foreign / export experience.
9. Creating bid for items from incorrect categories.

10. Reference of conditions published on any external site or reference to external documents/clauses.
11. Asking for any Tender fee / Bid Participation fee, as the case may be.
12. Buyer added ATC Clauses which are in contravention of clauses defined in bid detail section, including specifications, EMD Detail, ePBG Detail and MII and MSE Purchase Preference sections of the bid, unless otherwise allowed by the applicable GeM GTC.
13. Any ATC clause in contravention with GeM GTC Clause 4 (xiii) (h) will be invalid. In case of multiple L1 bidders against a service bid, the buyer shall place the Contract by selection of a bidder amongst the L-1 bidders through a Random Algorithm executed by GeM system.
14. In a category based bid, adding additional items, through buyer added, additional scope of work/ additional terms and conditions/or any other document. If buyer needs more items along with the main item, the same must be added through bunching category based items or by bunching custom catalogues or bunching a BoQ with the main category based item, the same must not be done through ATC or Scope of Work.

Further, if any seller has any objection/grievance against these additional clauses or otherwise on any aspect of this bid, they can raise their representation against the same by using the Representation window provided in the bid details field in Seller dashboard after logging in as a seller. Buyer is duty bound to reply to all such representations and would not be allowed to open bids if he fails to reply to such representations.

All GeM Sellers/Service Providers shall ensure full compliance with all applicable labour laws, including the provisions, rules, schemes and guidelines under the four Labour Codes i.e. the Code on Wages, 2019; the Industrial Relations Code, 2020; the Occupational Safety, Health and Working Conditions Code, 2020; and the Code on Social Security, 2020 as and when notified and brought into force by the Government of India.

For all provisions of the Labour Codes that are pending operationalisation through rules, schemes or notifications, the corresponding provisions of the pre-existing labour enactments (such as The Minimum Wages Act, 1948, The Payment of Wages Act, 1936, The Payment of Bonus Act, 1965, The Equal Remuneration Act, 1976, The Payment of Gratuity Act, 1972, etc. and relevant State Rules) shall continue to remain applicable.

The Seller/ Service Providers shall, therefore, be responsible for ensuring compliance under:

- **All notified and enforceable provisions of the new Labour Codes as mentioned hereinabove; and**
- **All operative provisions of the erstwhile Labour Laws until their complete substitution.**

All obligations relating to wages, social security, safety, working conditions, industrial relations etc. and any other statutory requirements shall be strictly met by the Seller/ Service Provider. Any non-compliance shall constitute a breach of the contract and shall entitle the Buyer to take appropriate action in accordance with the contract and applicable law.

This Bid is governed by the General Terms and Conditions, conditions stipulated in Bid and Service Level Agreement specific to the Service, as the case may be, as provided in the Marketplace.

However, in case of Service, if any condition specified in General Terms and Conditions is contradicted by the conditions stipulated in Service Level Agreement specific to said Service, then it will over-ride the conditions in the General Terms and Conditions.

[यह बिड सामान्य शर्तों के अंतर्गत भी शासित है /This Bid is also governed by the General Terms and Conditions](#)

जेम की सामान्य शर्तों के खंड 26 के संदर्भ में भारत के साथ भूमि सीमा साझा करने वाले देश के बिडर से खरीद पर प्रतिबंध के संबंध में भारत के साथ भूमि सीमा साझा करने वाले देश का कोई भी बिडर इस निविदा में बिड देने के लिए तभी पात्र होगा जब वह बिड देने वाला सक्षम प्राधिकारी के पास पंजीकृत हो। बिड में भाग लेते समय बिडर को इसका अनुपालन करना होगा और कोई भी गलत घोषणा किए जाने व इसका अनुपालन न करने पर अनुबंध को तत्काल समाप्त करने और कानून के अनुसार आगे की कानूनी कार्यवाई का आधार होगा।/In terms of GeM GTC clause 26 regarding Restrictions on procurement from a bidder of a country which shares a land border with India, any bidder from a country which shares a land border with India will be eligible to bid in this tender only if the bidder is registered with the Competent Authority. While participating in bid, Bidder has to undertake compliance of this and any false declaration and non-compliance of this would be a ground for immediate termination of the contract and further legal action in accordance with the laws.

---धन्यवाद/Thank You---

