

बिड दस्तावेज़ / Bid Document

| बिड विवरण/Bid Details | |
|---|--|
| बिड बंद होने की तारीख/समय /Bid End Date/Time | 02-02-2026 11:00:00 |
| बिड खुलने की तारीख/समय /Bid Opening Date/Time | 02-02-2026 11:30:00 |
| बिड पेशकश वैधता (बंद होने की तारीख से)/Bid Offer Validity (From End Date) | 180 (Days) |
| मंत्रालय/राज्य का नाम/Ministry/State Name | Pmo |
| विभाग का नाम/Department Name | Department Of Atomic Energy |
| संगठन का नाम/Organisation Name | Directorate Of Purchase And Stores |
| कार्यालय का नाम/Office Name | Directorate Of Purchase And Stores |
| कुल मात्रा/Total Quantity | 1 |
| वस्तु श्रेणी /Item Category | PET- CT 64 Slice Scanning System (Q2) |
| वर्षों के अनुभव एवं टर्नओवर से एमएसई को छूट प्राप्त है / MSE Relaxation for Years Of Experience and Turnover | Yes Complete |
| स्टार्टअप के लिए अनुभव के वर्षों और टर्नओवर से छूट प्रदान की गई है / Startup Relaxation for Years Of Experience and Turnover | Yes Complete |
| विक्रेता से मांगे गए दस्तावेज़/Document required from seller | Certificate (Requested in ATC) *In case any bidder is seeking exemption from Experience / Turnover Criteria, the supporting documents to prove his eligibility for exemption must be uploaded for evaluation by the buyer |
| क्या आप निविदाकारों द्वारा अपलोड किए गए दस्तावेज़ों को निविदा में भाग लेने वाले सभी निविदाकारों को दिखाना चाहते हैं? संदर्भ मेनू है/Do you want to show documents uploaded by bidders to all bidders participated in bid? | Yes (Documents submitted as part of a clarification or representation during the tender/bid process will also be displayed to other participated bidders after log in) |
| बिड लगाने की समय सीमा स्वतः नहीं बढ़ाने के लिए आवश्यक बिड की संख्या। / Minimum number of bids required to disable automatic bid extension | 1 |
| दिनों की संख्या, जिनके लिए बिड लगाने की समय-सीमा बढ़ाई जाएगी। / Number of days for which Bid would be auto-extended | 7 |

| बिड विवरण/Bid Details | |
|--|-----------------------------------|
| ऑटो एक्सटेंशन अधिकतम कितनी बार किया जाना है। / Number of Auto Extension count | 1 |
| बिड से रिवर्स नीलामी सक्रिय किया/Bid to RA enabled | Yes |
| रिवर्स नीलामी योग्यता नियम/RA Qualification Rule | H1-Highest Priced Bid Elimination |
| बिड का प्रकार/Type of Bid | Two Packet Bid |
| तकनीकी मूल्यांकन के दौरान तकनीकी स्पष्टीकरण हेतु अनुमत समय /Time allowed for Technical Clarifications during technical evaluation | 3 Days |
| निरीक्षण आवश्यक (सूचीबद्ध निरीक्षण प्राधिकरण /जेम के साथ पूर्व पंजीकृत एजेंसियों द्वारा)/Inspection Required (By Empanelled Inspection Authority / Agencies pre-registered with GeM) | No |
| मूल्यांकन पद्धति/Evaluation Method | Total value wise evaluation |
| वित्तीय दस्तावेज की आवश्यकता है / Financial Document Required | Yes |
| मध्यस्थता खंड/Arbitration Clause | No |
| सुलह खंड/Mediation Clause | No |

ईएमडी विवरण/EMD Detail

| | |
|-----------------------------|---------------------|
| एडवाइजरी बैंक/Advisory Bank | State Bank of India |
| ईएमडी राशि/EMD Amount | 2000000 |

ईपीबीजी विवरण /ePBG Detail

| | |
|---|---------------------|
| एडवाइजरी बैंक/Advisory Bank | State Bank of India |
| ईपीबीजी प्रतिशत (%) /ePBG Percentage(%) | 5.00 |
| ईपीबीजी की आवश्यक अवधि (माह) /Duration of ePBG required (Months). | 68 |

(a). जेम की शर्तों के अनुसार ईएमडी छूट के इच्छुक बिडर को संबंधित कैटेगरी के लिए बिड के साथ वैध समर्थित दस्तावेज प्रस्तुत करने है। एमएसई कैटेगरी के अंतर्गत केवल वस्तुओं के लिए विनिर्माता तथा सेवाओं के लिए सेवा प्रदाता ईएमडी से छूट के पात्र हैं। व्यापारियों को इस नीति के दायरे से बाहर रखा गया है।/EMD EXEMPTION: The bidder seeking EMD exemption, must submit the valid supporting document for the relevant category as per GeM GTC with the bid. Under MSE category, only manufacturers for goods and Service Providers for Services are eligible for exemption from EMD. Traders are excluded from the purview of this Policy.

(b). ईएमडी और संपादन जमानत राशि, जहां यह लागू होती है, लाभार्थी के पक्ष में होनी चाहिए। / EMD & Performance security should be in favour of Beneficiary, wherever it is applicable.

लाभार्थी /Beneficiary :

Acting through Director DPS, the President of India

Directorate Of Purchase And Stores, Department of Atomic Energy, Directorate of Purchase and Stores,
B1/5379/TPT
(.)

बोली विभाजन लागू नहीं किया गया/ Bid splitting not applied.

एमएसई खरीद वरीयता/MSE Purchase Preference

| | |
|---|----|
| एमएसई खरीद वरीयता/MSE Purchase Preference | No |
|---|----|

एमएसई के लिए सक्षम प्राधिकारी का विवरण:/Details of the Competent Authority for MSE

| | |
|--|--|
| सक्षम प्राधिकारी का नाम/Name of Competent Authority | Sher Bahadur |
| सक्षम प्राधिकारी का पदनाम/Designation of Competent Authority | Under Secretary (Procurement Policy) |
| सक्षम प्राधिकारी का कार्यालय/विभाग/प्रभाग/Office / Department / Division of Competent Authority | Department of Expenditure |
| सीए अनुमोदन संख्या/CA Approval Number | No.F.4/1/2023-PPD(pt.) |
| सक्षम प्राधिकारी अनुमोदन तिथि/Competent Authority Approval Date | 28-06-2024 |
| सक्षम प्राधिकारी द्वारा प्रदान की गई स्वीकृति का संक्षिप्त विवरण//Brief Description of the Approval Granted by Competent Authority | Available at Item no 281 of GTE exemption List |

Competent Authority Approval for not opting Micro and Small Enterprises Preference : [View Document](#)

एमआईआई खरीद वरीयता/MII Purchase Preference

| | |
|--|----|
| एमआईआई खरीद वरीयता/MII Purchase Preference | No |
|--|----|

एमआईआई के लिए सक्षम प्राधिकारी का विवरण:/Details of the Competent Authority for MII

| | |
|---|--|
| सक्षम प्राधिकारी का नाम/Name of Competent Authority | Sher Bahadur |
| सक्षम प्राधिकारी का पदनाम/Designation of Competent Authority | Under Secretary (Procurement Policy) |
| सक्षम प्राधिकारी का कार्यालय/विभाग/प्रभाग/Office / Department / Division of Competent Authority | Department of Expenditure |
| सीए अनुमोदन संख्या/CA Approval Number | No.F.4/1/2023-PPD(pt.) |
| सक्षम प्राधिकारी अनुमोदन तिथि/Competent Authority Approval Date | 28-06-2024 |
| सक्षम प्राधिकारी द्वारा प्रदान की गई स्वीकृति का संक्षिप्त विवरण/Brief Description of the Approval Granted by Competent Authority | Available at Item no 281 of GTE exemption List |

Competent Authority Approval for not opting Make In India Preference : [View Document](#)

1. If the bidder is a Micro or Small Enterprise as per latest orders issued by Ministry of MSME, the bidder shall be relaxed from the eligibility criteria of "Experience Criteria" as defined above subject to meeting of quality and technical specifications. The bidder seeking Relaxation from Experience Criteria, shall upload the supporting documents to prove his eligibility for Relaxation.
2. If the bidder is a Micro or Small Enterprise (MSE) as per latest orders issued by Ministry of MSME, the bidder shall be relaxed from the eligibility criteria of "Bidder Turnover" as defined above subject to meeting of quality and technical specifications. If the bidder itself is MSE OEM of the offered products, it would be relaxed from the "OEM Average Turnover" criteria also subject to meeting of quality and technical specifications. The bidder seeking Relaxation from Turnover, shall upload the supporting documents to prove his eligibility for Relaxation.
3. If the bidder is a DPIIT registered Startup, the bidder shall be relaxed from the the eligibility criteria of "Experience Criteria" as defined above subject to their meeting of quality and technical specifications. The bidder seeking Relaxation from Experience Criteria, shall upload the supporting documents to prove his eligibility for Relaxation.
4. If the bidder is a DPIIT registered Startup, the bidder shall be relaxed from the the eligibility criteria of "Bidder Turnover" as defined above subject to their meeting of quality and technical specifications. If the bidder is DPIIT Registered OEM of the offered products, it would be relaxed from the "OEM Average Turnover" criteria also subject to meeting of quality and technical specifications. The bidder seeking Relaxation from Turnover shall upload the supporting documents to prove his eligibility for Relaxation.
5. Estimated Bid Value indicated above is being declared solely for the purpose of guidance on EMD amount and for determining the Eligibility Criteria related to Turn Over, Past Performance and Project / Past Experience etc. This has no relevance or bearing on the price to be quoted by the bidders and is also not going to have any impact on bid participation. Also this is not going to be used as a criteria in determining reasonableness of quoted prices which would be determined by the buyer based on its own assessment of reasonableness and based on competitive prices received in Bid / RA process.
6. Reverse Auction would be conducted amongst all the technically qualified bidders except the Highest quoting bidder. The technically qualified Highest Quoting bidder will not be allowed to participate in RA. However, H-1 will also be allowed to participate in RA in following cases:
 - i. If number of technically qualified bidders are only 2 or 3.
 - ii. If Buyer has chosen to split the bid amongst N sellers, and H1 bid is coming within N.
 - iii. In case Primary product of only one OEM is left in contention for participation in RA on elimination of H-1.
 - iv. If L-1 is non-MSE and H-1 is eligible MSE and H-1 price is coming within price band of 15% of Non-MSE L-1
 - v. If L-1 is non-MII and H-1 is eligible MII and H-1 price is coming within price band of 20% of Non-MII L-1

एक्सेल में अपलोड किए जाने की आवश्यकता /Excel Upload Required :

Charges for CMC - [1764669549.xlsx](#)

Pre Bid Detail(s)

| मूल्य भिन्नता खंड दस्तावेज़/Pre-Bid Date and Time | प्री-बिड स्थान/Pre-Bid Venue |
|---|---|
| 12-01-2026 11:00:00 | Room No. 411, 4th Floor, RMC, TMC Annexe, TMH, Parel, Mumbai-400012 |

PET- CT 64 Slice Scanning System (1 pieces)

तकनीकी विशिष्टियाँ /Technical Specifications

* जेम केटेगरी विशिष्टि के अनुसार / As per GeM Category Specification

| विवरण/Specification | विशिष्टि का नाम /Specification Name | बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values) |
|---------------------|-------------------------------------|---|
| DETECTOR ASSEMBLY | Crystal material | LSO |
| | Crystal size (mm) | 4 x 4 x 20 |
| X-RAY TUBE | Tube cooling technique | Oil cooling |

| विवरण/Specification | विशिष्टि का नाम /Specification Name | बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values) |
|--|---|---|
| WORKSTATION SPECIFICATIONS | OEM post processing workstation with dual medical grade Monitor | Server |
| | If it is server workstation, number of nodes to be provided | 3 |
| | Number of Concurrent licenses/users to be provided | 2 |
| SERVER HARDWARE | Image storage capacity of Server | ? 2 TB |
| | Ram Size of Server | ? 32 GB |
| | Monitor Size of Server | 19 inch or more |
| | Monitor Type of Server | LCD/LED/TFT |
| | Minimum rendering capacity of server | 20,000 Concurrent Slices |
| WORKSTATION HARDWARE | Minimum Ram Size of Workstation | 16 Or higher (gigabyte) |
| | Minimum Storage capacity of Workstation | 1000 Or higher (gigabyte) |
| CLINICAL APPLICATION SOFTWARE ON WORKSTATION | Dynamic PET processing and viewing software | Yes |
| | Software for PET/CT/MRI Fusion | Yes |
| | Multi modality image registration software | Yes |
| | Provision of make DICOM/PDF/JPEG/AVI/MPEG digital output | Yes |
| | Neuro quantification software including assessment of dementia using SUV values | Yes |
| | Complete cardiac package both for PET and CT with ECG gated studies and ECG gated dose modulation | Yes |
| | Software to evaluate different scans of same patient done over a time to give overall assessment of malignancies progression in oncology | Yes |
| | PET DICOM 3 or higher version facilities for clinical applications and ability to import MR/CT DICOM data | Yes |

| विवरण/Specification | विशिष्टि का नाम /Specification Name | बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values) |
|---------------------|--|---|
| | Software for data collection, CT based attenuation correction, reconstruction of image for co registration | Yes |
| | Full 3-D prospective reconstruction with iterative scatter correction, advance 3-D Volume rendering with 3-D fusion, virtual endoscopy, Model / image based 3-D scatter correction, MIP, whole body acquisition, dynamic acquisition | Yes |
| | Remote service diagnostic facility | Yes |
| ACCESSORIES | Online UPS for full backup of atleast 30 minutes for the entire system | Yes |
| | Patient table accessories | Yes |
| | One Compatible Color Printer of standard make | Yes |
| | Variety of phantoms for quality control | Yes |
| | One Dual Head Pressure injector compatible with CT and 200 disposable CT syringes with tubing and connector | Yes, No |
| | One Single Head Pressure injector with 50 syringes and tubing sets | Yes, No |
| | Contrast Injection Integration with Injector | Yes, No |
| | ECG gating device and necessary electronics to enable gated cardiac acquisition | Yes |
| | One dose calibrator for PET radionuclides | Yes |
| | Two Vein Finder | Yes, No |
| | One Lead Mobile Shield Barrier of 4 Feet Height | No |

| विवरण/Specification | विशिष्टि का नाम /Specification Name | बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values) |
|-----------------------------------|---|---|
| | Five Thyroid Lead Shields to protect CT radiation Exposure to Thyroid to person performing Guided Procedures | No |
| CERTIFICATIONS AND REPORTS | Product certification | EU-CE, US-FDA |
| | Conformity to Manufacturer's Certification | ISO 9001 & ISO 13485 |
| PRE-INSTALLATION SITE REQUIREMENT | Preparation of PET CT Scan Room , console Room , Technical room, wash rooms, dose administration room, hot lab, post injection etc to be done by the seller in the area of approx 1700 sq feet | Yes |
| | Tile based false ceiling in the rooms to be done by the seller | Yes |
| | Vinyl flooring in all the rooms to be done by the seller | Yes |
| | Necessary modification civil work, electrical , plumbing, furnishing, air conditioning in the given area to be done by the seller | Yes |
| | Room shielding to be done by the seller | As Per AERB Guidelines |
| | Any other necessary work for system functioning like earthing , proper LED lighting power points etc to be done by the seller | Yes |
| INSTALLATION & TRAINING | Number of installations of the PET CT in Central /State/PSU Govt Hospitals(Note: Seller should supply a performance certificate of the device to the buyer) | upto 3 |
| WARRANTY | Warranty in years (Option of comprehensive warranty is available through bidding only, which if opted will supersede normal warranty in the catalogue) | 5 Or higher |

Additional Specification Parameters - PET- CT 64 Slice Scanning System (1 pieces)

| Specification Parameter Name | Bid Requirement (Allowed Values) |
|--|---|
| 1- PET Detector Assembly 2- CT SPECIFICATION | (a) Crystal material- LSO /LYSO/LBS, (b) Crystal size (mm)- 4x4x20, (c) Axial FOV (cm)- More than or equal to 24 2- (a) Gantry Aperture- More than or equal to 70 cm, (b) Number of slices acquired simultaneously- More than or equal to 64, (c) Field of view CTAC- More than or equal to 70 cm |
| Performance Parameters | (a)All specifications must comply with latest NEMA standards publication NU2 performance measurements- Yes,(b)System sensitivity- More than or equal to 12 cps/kBq (3D), (c)Peak NECR-More than or equal to 85 kcps, (d)Axial and Transaxial spatial resolution at 1 cm and 10 cm from the central axis of the gantry - Less than or equal to 5mm, (e)Energy resolution (%)- Less than or equal to 10,(f) Timing resolution (psec)- Less than or equal to 500, (g) SiPM enabled photon detection technology |
| 1- X-RAY Assembly 2- Patient Couch 3- Operator Console with Image display and management | 1- (a)Max Power Output-70 kW or more, (b)Max Tube Voltage-140 kV, (c)Max Tube current-600 or more (d)Dose modulation technique-Yes, (e)Cardiac specific dose saving technique-Yes, (f)Anode Heat storage (MHU)- More than or equal to 7 2-(a) Table Load capacity-200kg or more, (b)Horizontal scan range-200 cm or more 3-(a)Fully integrated PET CT console-1, (b)App for reduction Vol. & metal artifact-Yes,(c)Filming /Printing capability-Yes |
| 1-Image Processing Software 2- Motion Management 3- Quantitative Analysis for both RECIST and PERCIST criteria 4- Dual Head Pressure injector 5- UPS 6-Integration with HIS & PACS 7- Contrast injection integration | 1-(a)Oncology Application with independent Lifetime License on all workstations-Yes (b)Complete PET Cardiac package with normal database - Yes,(c)Neuro quantification software with normal database-Yes 2-Motion management with CT attenuation correction software-Yes, 3-Yes 4-Yes, One set with connector and tubing sets 5-Yes, with isolation transformer 6-Yes, will be bidder's responsibility 7- Yes, ONE set |
| 1- Installation and Commissioning 2- CMC after Warranty period | 1- Dismantling of old PET/CT system and transport to BARC, Trombay and site preparation as per AERB requirement. 2- CMC budgetary quotation for 5 years after warranty period. |

* Bidders offering must also comply with the additional specification parameters mentioned above.

उपरोक्त मद के लिए इंस्टॉलेशन कमीशनिंग एंड टेस्टिंग (आईसीटी) का विवरण / Installation Commissioning and Testing (ICT) details for the above item:

| | |
|---|---------|
| % of Product Cost Payable on Product Delivery | 80% |
| Min Cost Allocation for ICT as a % of product cost | 20% |
| Number of days allowed for ICT after site readiness communication to seller | 45 Days |

परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

| क्र.सं./S.N o. | परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer | पता/Address | मात्रा /Quantity | डिलीवरी के दिन/Delivery Days |
|----------------|---|-------------|------------------|------------------------------|
| | | | | |

| क्र.सं./S.N o. | परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer | पता/Address | मात्रा /Quantity | डिलीवरी के दिन/Delivery Days |
|-------------------|---|---|------------------|---------------------------------|
| 1 | Manoj Bhargaw Golatkar | 400094,DIRECTORATE OF PURCHASE AND STORES V. S. BHAVAN ANUSHAKTINAGAR | 1 | 180 |

Special terms and conditions-Version:2 effective from 21-10-2022 for category PET- CT 64 Slice Scanning System

1. 1. All Provisions of Drugs and Cosmetics Act, 1940 and Rules made there under as amended till date will always be applicable. This will include all notifications issued by *Central Drugs Standard Control Organisation(CDSCO)*, Ministry of Health & Family Welfare (MoHFW) and Department of Pharmaceuticals (DOP), Ministry of Chemicals & Fertilizers time to time in this regard.
2. The sellers are registered on GeM based on self-declaration of valid Drug License, product certification, test reports etc. However, buyers must check and validate the details at their end for all applicable licenses and certifications e.g., validity and authenticity/genuineness of drug license, product certification, manufacturer certification/licenses, test reports etc.
3. The price offered by the seller shall not, in any case exceed the DPCO controlled price, if any, fixed by the Central/State Government, the Maximum Retail Price (MRP) and the selling price. The seller must reduce the prices if there is any reduction in DPCO ceiling price, if any.
4. **Comprehensive warranty:** Comprehensive warranty shall include preventive maintenance including calibration as per technical/ service /operational manual of the manufacturer, service charges and spares. During the warranty period commencing from date of the successful completion of warranty period, Service personnel shall visit each consignee site as recommended in the manufacturer's technical/ service /operational manual, at least once in six months. warranty shall not be including the consumables. Further there will be 98% uptime warranty during warranty period on 24 (hrs) X 7 (days) X 365 (days) basis, with penalty, to extend warranty period by double the downtime period.
5. **Service centres:** Details of Service outlets in India to render services for equipment to be furnished to buyer/consignees with complete address,telephone numbers, e mails etc at time of making the supplies. It shall be the responsibility of seller to ensure that authorized service centres are available to cater to the areas where supplies are made within reasonable distance from where the service calls can be handled.Details of toll-free numbers for service call and online registration of service requests also to be provided buyer/consignee at the time of supplies.
6. **Source of supply:** It shall be responsibility of seller to provide Documents regarding source of equipments such as copy of Performa invoice or any other documents to establish that the products supplied are manufactured by OEM indicated and sourced from them.
7. **Packing and Marking:** Medical equipments being very delicate and sensitive packing for the goods should be strong and durable enough to withstand transit including transshipment (if any), rough handling, open storage etc. without any damage, deterioration etc. .The size, weights and volumes of the packing cases, remoteness of the final destination of the goods, availability or otherwise of transport and handling facilities at all points during transit up to final destination,. Quality of packing, the manner of marking within & outside the packages and provision of accompanying documentation shall take into consideration the type of medical equipments being supplied. The accessories shall be suitably labelled and packed. Each of the package shall be marked on three sides with indelible paint of proper quality: indicating contract number and date, brief description of goods including quantity, Packing list reference number, country of origin of goods and any other relevant details.
8. **Spare Parts:** Seller shall provide materials, information etc. pertaining to spare parts manufactured and supplied by the OEM. It shall be ensured that the required spares are available for purchase at least for 10 years from date of supplies. In case due to any reasons the production of the spare parts is discontinued sufficient advance notice should be given to the buyer/consignee before such discontinuation to provide adequate time to purchase the required spare parts etc. Further, OEM and their service centres/dealers shall carry sufficient inventories to assure ex-stock supply of consumables and spares for the equipments so that the same are available.
OEM or reseller shall always accord most favoured client status to the buyer/consignee and shall give the most competitive price for spares and consumables of its machines/equipments supplied.
9. **Installation, Training, Manuals:** Seller shall be responsible to carry out Installation & commissioning, Supervision and Demonstration of the goods. They shall provide required jigs and

tools for assembly, minor civil works for the completion of the installation and Training of Consignee's representatives for operating and maintaining the equipment and supplying required number of operation & maintenance manual for the goods. In case the category parameters are specifying any requirements regarding the installations, training and manuals the same shall also be applicable.

10. **Electrical safety checking:** Sellers are required to make sure that they furnish the list of equipments for carrying out routine and preventive maintenance to buyer/consignee .They should make sure to periodically check the electrical safety aspects as per BIS Safety Standards or equivalent .In case they do not have required equipment for such testing should ensure that the equipments checked for electrical safety compliance through labs with facilities for such checking during every preventive maintenance call.
11. **Software:** All software updates should be provided free of cost during warranty period.
12. Any other Terms and Conditions which is not included or at variance with the conditions specified in STC/GTC, may be added by the buyer through Additional Terms and Conditions in the bid to ensure items are procured from authentic/validated source with appropriate and applicable quality. The above terms and conditions are in reverse order of precedence i.e. ATC shall supersede specific STC which shall supersede General Terms and Conditions ("GTC"), whenever there are any conflicting provisions.

क्रेता द्वारा जोड़ी गई बिड की विशेष शर्तें/**Buyer Added Bid Specific Terms and Conditions**

1. **Scope of Supply**

Scope of supply (Bid price to include all cost components) : Supply Installation Testing Commissioning of Goods and Training of operators and providing Statutory Clearances required (if any)

2. **Warranty**

Warranty period of the supplied products shall be 5 years from the date of final acceptance of goods or after completion of installation, commissioning & testing of goods (if included in the scope of supply), at consignee location. OEM Warranty certificates must be submitted by Successful Bidder at the time of delivery of Goods. The seller should guarantee the rectification of goods in case of any break down during the guarantee period. Seller should have well established Installation, Commissioning, Training, Troubleshooting and Maintenance Service group in INDIA for attending the after sales service. Details of Service Centres near consignee destinations are to be uploaded along with the bid.

3. **Forms of EMD and PBG**

Bidders can also submit the EMD with Account Payee Demand Draft in favour of

PAY & ACCOUNTS OFFICER, DPS, MUMBAI
payable at
MUMBAI

. Bidder has to upload scanned copy / proof of the DD along with bid and has to ensure delivery of hardcopy to the Buyer within 5 days of Bid End date / Bid Opening date.

4. **Forms of EMD and PBG**

Successful Bidder can submit the Performance Security in the form of Account Payee Demand Draft also (besides PBG which is allowed as per GeM GTC). DD should be made in favour of

PAY & ACCOUNTS OFFICER, DPS, MUMBAI
payable at
MUMBAI

. After award of contract, Successful Bidder can upload scanned copy of the DD in place of PBG and has to ensure delivery of hard copy to the original DD to the Buyer within 15 days of award of contract.

5. **Generic**

Scope of supply includes Training: Number of employees to be trained

10
, Place for Training
Purchaser's premises
and Duration of training
15
days.

6. Buyer Added Bid Specific ATC

Buyer Added text based ATC clauses

1. OUR REF:- DPS/CPU/04/B1/5379/TPT, WHICH MAY BE MENTIONED IN ALL CORRESPONDENCE.

2 THE PERSON/TRANSPORT CONTRACTOR WHO WILL BE DELIVERING THE MATERIAL SHOULD INVARIABLY CALL ON/REPORT TO THE STORE KEEPER IN CHARGE, FOR APPROPRIATE DECISION.

3. VEHICLES WITH GPS FACILITY WILL NOT BE ALLOWED ENTRY INSIDE PURCHASER'S PREMISES. IN VIEW OF THE ABOVE, THE CONSIGNMENTS SHOULD BE DISPATCHED ONLY THROUGH VEHICLES WITHOUT GPS SYSTEM.

4. The items covered by this contract shall be delivered to the consignee on all working days i.e. from Monday to Friday between 11.00 hrs to 12.45 hrs and again from 14.00 hrs to 17.00 hrs.

5. ANNEXURE-X : THE CONTRACTOR IS REQUIRED TO MAINTAIN HINDRANCE REGISTER FOR REPORTING HINDRANCES IF ANY WHILE EXECUTING THE WORK AS PER ATTACHED ANNEXURE-X. THE CONTRACTOR SHALL GET RECORD OF HINDRANCES IN THE HINDRANCE REGISTER(S) APPROVED/ENDORSED BY THE PURCHASER. SUCH HINDRANCE IN THE WORK ENDORSED BY THE PURCHASER WILL ONLY BE TAKEN INTO CONSIDERATION FOR GRANTING OF DELIVERY DATE REFIXATION/EXTENSION.

6. BIDDERS CAN ALSO MAKE THE PAYMENT OF EARNEST MONEY DEPOSIT (EMD) / PERFORMANCE SECURITY DEPOSIT BANK GUARANTEE (PSDBG) ELECTRONICALLY USING CREDIT CARD (CC), DEBIT CARD (DC), NET BANKING, IMMEDIATE PAYMENT SERVICE (IMPS) AND UNIFIED PAYMENT INTERFACE/BHARAT INTERFACE FOR MONEY (UPI/BHIM) THROUGH Non-Tax Receipt Portal (NTRP) known as Bharatkosh.

URL: <https://bharatkosh.gov.in>

THE DEPOSITOR/USER IS REQUESTED TO FOLLOW THE STEPS AS INDICATED IN ANNEXURE-I FOR MAKING PAYMENT. THE SOFT COPY OF THE ONLINE PAYMENT RECEIPT AND CHALLAN GENERATED THROUGH THE SYSTEM SHALL BE UPLOADED ALONG WITH THE BID AS PER THE FORM AT ENCLOSED HEREWITH FOR ENABLING REFUND OF EMD. IT MAY BE NOTED THAT IN THE "REMARKS" COLUMN,

THE REASONS FOR DEPOSITING THE MONEY INCLUDING TENDER NO. /BID REFERENCE / PURCHASE ORDER NO., REMITTER'S NAME, ADDRESS MAY BE MANDATORILY INDICATED AS EXPLAINED IN NOTE BELOW ANNEXURE-I, SO AS TO RELATE THE PAYMENTS WITH THE ACTIVITIES.

FLOW CHART FOR MAKING PAYMENT IN "BHARATKOSH" IS ATTACHED AS ANNEXURE-I.

Note : The Bank Guarantee/Demand draft should be forwarded to the address as mentioned :- Assistant Purchase Officer, Group : B1 Directorate of Purchase & Stores V S Bhavan, Anushaktinagar, Mumbai-400094 Email : apob1@dpsdae.gov.in

7. This item falls under medical devices list at serial number 281 in GTE exemption list.

ACTUAL PLACE OF DELIVERY

RMC Stores, 4th Floor, Room No.-425, Anex Bldg., Tata Memorial Hospital, Jeraba i Wadia road, Parel, Mumbai- 400012.

Charges for CMC to be quoted in the excel sheet provided with the tender. It is T O BE SUBMITTED AS A SEPARATE ATTACHMENT IN PRICE BID ONLY. CMC charges shall not be considered for evaluation.

KINDLY READ FOLLOWINGS AS:-

Image storage capacity of

Server is greater than or equal to 2 TB

RAM Size of Server is greater than or equal to 32 GB

7. Buyer Added Bid Specific ATC

Buyer uploaded ATC document [Click here to view the file.](#)

अस्वीकरण/Disclaimer

The additional terms and conditions have been incorporated by the Buyer after approval of the Competent Authority in Buyer Organization, whereby Buyer organization is solely responsible for the impact of these clauses on the bidding process, its outcome, and consequences thereof including any eccentricity / restriction arising in the bidding process due to these ATCs and due to modification of technical specifications and / or terms and conditions governing the bid. If any clause(s) is / are incorporated by the Buyer regarding following, the bid and resultant contracts shall be treated as null and void and such bids may be cancelled by GeM at any stage of bidding process without any notice:-

1. Definition of Class I and Class II suppliers in the bid not in line with the extant Order / Office Memorandum issued by DPIIT in this regard.
2. Seeking EMD submission from bidder(s), including via Additional Terms & Conditions, in contravention to exemption provided to such sellers under GeM GTC.
3. Publishing Custom / BOQ bids for items for which regular GeM categories are available without any Category item bunched with it.
4. Creating BoQ bid for single item.
5. Mentioning specific Brand or Make or Model or Manufacturer or Dealer name.
6. Mandating submission of documents in physical form as a pre-requisite to qualify bidders.
7. Floating / creation of work contracts as Custom Bids in Services.
8. Seeking sample with bid or approval of samples during bid evaluation process. (However, in bids for [attached categories](#), trials are allowed as per approved procurement policy of the buyer nodal Ministries)
9. Mandating foreign / international certifications even in case of existence of Indian Standards without specifying equivalent Indian Certification / standards.
10. Seeking experience from specific organization / department / institute only or from foreign / export experience.
11. Creating bid for items from irrelevant categories.
12. Incorporating any clause against the MSME policy and Preference to Make in India Policy.
13. Reference of conditions published on any external site or reference to external documents/clauses.
14. Asking for any Tender fee / Bid Participation fee / Auction fee in case of Bids / Forward Auction, as the case may be.
15. Buyer added ATC Clauses which are in contravention of clauses defined by buyer in system generated bid template as indicated above in the Bid Details section, EMD Detail, ePBG Detail and MII and MSE Purchase

Preference sections of the bid, unless otherwise allowed by GeM GTC.

16. In a category based bid, adding additional items, through buyer added additional scope of work/ additional terms and conditions/or any other document. If buyer needs more items along with the main item, the same must be added through bunching category based items or by bunching custom catalogs or bunching a BoQ with the main category based item, the same must not be done through ATC or Scope of Work.

Further, if any seller has any objection/grievance against these additional clauses or otherwise on any aspect of this bid, they can raise their representation against the same by using the Representation window provided in the bid details field in Seller dashboard after logging in as a seller within 4 days of bid publication on GeM. Buyer is duty bound to reply to all such representations and would not be allowed to open bids if he fails to reply to such representations.

All GeM Sellers/Service Providers shall ensure full compliance with all applicable labour laws, including the provisions, rules, schemes and guidelines under the four Labour Codes i.e. the Code on Wages, 2019; the Industrial Relations Code, 2020; the Occupational Safety, Health and Working Conditions Code, 2020; and the Code on Social Security, 2020 as and when notified and brought into force by the Government of India.

For all provisions of the Labour Codes that are pending operationalisation through rules, schemes or notifications, the corresponding provisions of the pre-existing labour enactments (such as The Minimum Wages Act, 1948, The Payment of Wages Act, 1936, The Payment of Bonus Act, 1965, The Equal Remuneration Act, 1976, The Payment of Gratuity Act, 1972, etc. and relevant State Rules) shall continue to remain applicable.

The Seller/ Service Providers shall, therefore, be responsible for ensuring compliance under:

- **All notified and enforceable provisions of the new Labour Codes as mentioned hereinabove; and**
- **All operative provisions of the erstwhile Labour Laws until their complete substitution.**

All obligations relating to wages, social security, safety, working conditions, industrial relations etc. and any other statutory requirements shall be strictly met by the Seller/ Service Provider. Any non-compliance shall constitute a breach of the contract and shall entitle the Buyer to take appropriate action in accordance with the contract and applicable law.

[यह बिड सामान्य शर्तों के अंतर्गत भी शासित है /This Bid is also governed by the General Terms and Conditions](#)

जेम की सामान्य शर्तों के खंड 26 के संदर्भ में भारत के साथ भूमि सीमा साझा करने वाले देश के बिडर से खरीद पर प्रतिबंध के संबंध में भारत के साथ भूमि सीमा साझा करने वाले देश का कोई भी बिडर इस निविदा में बिड देने के लिए तभी पात्र होगा जब वह बिड देने वाला सक्षम प्राधिकारी के पास पंजीकृत हो। बिड में भाग लेते समय बिडर को इसका अनुपालन करना होगा और कोई भी गलत घोषणा किए जाने व इसका अनुपालन न करने पर अनुबंध को तत्काल समाप्त करने और कानून के अनुसार आगे की कानूनी कार्यवाई का आधार होगा।/In terms of GeM GTC clause 26 regarding Restrictions on procurement from a bidder of a country which shares a land border with India, any bidder from a country which shares a land border with India will be eligible to bid in this tender only if the bidder is registered with the Competent Authority. While participating in bid, Bidder has to undertake compliance of this and any false declaration and non-compliance of this would be a ground for immediate termination of the contract and further legal action in accordance with the laws.

---धन्यवाद/Thank You---