

बिड दस्तावेज़ / Bid Document

बिड विवरण / Bid Details	
बिड बंद होने की तारीख/समय / Bid End Date/Time	28-03-2026 14:00:00
बिड खुलने की तारीख/समय / Bid Opening Date/Time	28-03-2026 14:30:00
बिड पेशकश वैधता (बंद होने की तारीख से) / Bid Offer Validity (From End Date)	30 (Days)
मंत्रालय/राज्य का नाम / Ministry/State Name	Rajasthan
विभाग का नाम / Department Name	Department Of Higher Education
संगठन का नाम / Organisation Name	Jai Narayan Vyas University Jodhpur
कार्यालय का नाम / Office Name	Jai Narain Vyas University Jodhpur
कुल मात्रा / Total Quantity	2
वस्तु श्रेणी / Item Category	Ultra low temperature laboratory deep freezer (Q3)
बिडर का न्यूनतम औसत वार्षिक टर्नओवर (3 वर्षों का) / Minimum Average Annual Turnover of the bidder (For 3 Years)	17 Lakh (s)
मूल उपकरण निर्माता का औसत टर्नओवर (गत 3 वर्षों का) / OEM Average Turnover (Last 3 Years)	134 Lakh (s)
उन्हीं/समान सेवा के लिए अपेक्षित विगत अनुभव के वर्ष / Years of Past Experience Required for same/similar service	3 Year (s)
एमएसएमई के लिए अनुभव के वर्षों और टर्नओवर से छूट प्रदान की गई है / MSE Relaxation for Years of Experience and Turnover	No
स्टार्टअप के लिए अनुभव के वर्षों और टर्नओवर से छूट प्रदान की गई है / Startup Relaxation for Years of Experience and Turnover	No
विक्रेता से मांगे गए दस्तावेज़ / Document required from seller	Experience Criteria, Past Performance, Bidder Turnover, Certificate (Requested in ATC), OEM Authorization Certificate, OEM Annual Turnover, Additional Doc 1 (Requester ATC), Compliance of BoQ specification and supporting documents. *In case any bidder is seeking exemption from Experience / Turnover Criteria, the supporting documents to prove his eligibility for exemption must be uploaded for evaluation by the buyer

बिड विवरण/Bid Details

क्या आप निविदाकारों द्वारा अपलोड किए गए दस्तावेजों को निविदा में भाग लेने वाले सभी निविदाकारों को दिखाना चाहते हैं? संदर्भ मेन् है/Do you want to show documents uploaded by bidders to all bidders participated in bid?	Yes (Documents submitted as part of a clarification or representation during the tender/bid process will also be displayed to other participated bidders after log in)
बिड लगाने की समय सीमा स्वतः नहीं बढ़ाने के लिए आवश्यक बिड की संख्या। / Minimum number of bids required to disable automatic bid extension	1
दिनों की संख्या, जिनके लिए बिड लगाने की समय-सीमा बढ़ाई जाएगी। / Number of days for which Bid would be auto-extended	3
ऑटो एक्सटेंशन अधिकतम कितनी बार किया जाना है। / Number of Auto Extension count	1
विगत प्रदर्शन /Past Performance	80 %
बिड से रिवर्स नीलामी सक्रिय किया/Bid to RA enabled	No
बिड का प्रकार/Type of Bid	Two Packet Bid
तकनीकी मूल्यांकन के दौरान तकनीकी स्पष्टीकरण हेतु अनुमत समय /Time allowed for Technical Clarifications during technical evaluation	3 Days
निरीक्षण आवश्यक (सूचीबद्ध निरीक्षण प्राधिकरण /जेम के साथ पूर्व पंजीकृत एजेंसियों द्वारा)/Inspection Required (By Empanelled Inspection Authority / Agencies pre-registered with GeM)	No
Payment Timelines	Payments shall be made to the Seller within 45 days of issue consignee receipt-cum-acceptance certificate (CRAC) and on-line submission of bills (This is in supersession of 10 days time as provided in clause 12 of GeM GTC)
मूल्यांकन पद्धति/Evaluation Method	Total value wise evaluation
मध्यस्थता खंड/Arbitration Clause	No
सुलह खंड/Mediation Clause	No

ईएमडी विवरण/EMD Detail

एडवाइजरी बैंक/Advisory Bank	Bank Of Baroda
ईएमडी राशि/EMD Amount	66800

ईपीबीजी विवरण /ePBG Detail

एडवाइजरी बैंक/Advisory Bank	Bank Of Baroda
ईपीबीजी प्रतिशत (%) /ePBG Percentage(%)	5.00

ईपीबीजी की आवश्यक अवधि (माह) /Duration of ePBG required (Months).

69

(a). जेम की शर्तों के अनुसार ईएमडी छूट के इच्छुक बिडर को संबंधित कटेगरी के लिए बिड के साथ वैध समर्थित दस्तावेज प्रस्तुत करने है। एमएसई केटेगरी अंतर्गत केवल वस्तुओं के लिए विनिर्माता तथा सेवाओं के लिए सेवा प्रदाता ईएमडी से छूट के पात्र हैं। व्यापारियों को इस नीति के दायरे से बाहर रखा गया है।/EMD EXEMPTION: The bidder seeking EMD exemption, must submit the valid supporting document for the relevant category as per GeM GTC with the bid. Under MSE category, only manufacturers for goods and Service Providers for Services are eligible for exemption from EMD. Traders are excluded from the purview of this Policy.

(b). ईएमडी और संपादन जमानत राशि, जहां यह लागू होती है, लाभार्थी के पक्ष में होनी चाहिए। / EMD & Performance security should be in favour of Beneficiary, wherever it is applicable.

लाभार्थी /Beneficiary :

Registrar

Jai Narain Vyas University Jodhpur, DEPARTMENT OF HIGHER EDUCATION, Jai Narayan Vyas University Jodhpur, (Sandeep Sandoo)

बोली विभाजन लागू नहीं किया गया/ Bid splitting not applied.

एमएसई खरीद वरीयता/MSE Purchase Preference

एमएसई खरीद वरीयता/MSE Purchase Preference	No
---	----

एमआईआई खरीद वरीयता/MII Purchase Preference

एमआईआई खरीद वरीयता/MII Purchase Preference	No
--	----

1. The minimum average annual financial turnover of the bidder during the last three years, ending on 31st March of previous financial year, should be as indicated above in the bid document. Documentary evidence in the form of certified Audited Balance Sheets of relevant periods or a certificate from the Chartered Accountant / Cost Accountant indicating the turnover details for the relevant period shall be uploaded with the bid. In case the date of constitution / incorporation of the bidder is less than 3-year-old, the average turnover in respect of the completed financial years after the date of constitution shall be taken into account for this criteria.

2. Experience Criteria: In respect of the filter applied for experience criteria, the Bidder or its OEM of the product offered in the bid {themselves or through reseller(s)} should have regularly, manufactured and supplied same or similar Category Products to any Central / State Govt Organization / PSU for number of Financial years as indicated above in bid document before the bid opening date. Copies of relevant contracts and delivery acceptance certificates like CR/ to be submitted along with bid in support of having supplied some quantity during each of the Financial year. In case of bunch bids, the category of primary product having highest value should meet this criterion.

3. OEM Turn Over Criteria: The minimum average annual financial turnover of the OEM of the offered product during last three years, ending on 31st March of the previous financial year, should be as indicated in the bid document. Documentary evidence in the form of certified Audited Balance Sheets of relevant periods or a certificate from the Chartered Accountant / Cost Accountant indicating the turnover details for the relevant period shall be uploaded with bid. In case the date of constitution / incorporation of the OEM is less than 3 year old, the average turnover in respect of the completed financial years after the date of constitution shall be taken into account for this criteria.

4. Past Performance: The Bidder or its OEM {themselves or through re-seller(s)} should have supplied same or similar Category Products for 80% of bid quantity, in at least one of the last three Financial years before the bid opening date to any Central / State Govt Organization / PSU. Copies of relevant contracts (proving supply of cumulative order quantity in any one financial year) to be submitted along with bid in support of quantity supplied in the relevant Financial year. In case of bunch bids, the category related to primary product having highest bid value should meet this criterion.

Ultra Low Temperature Laboratory Deep Freezer (2 pieces)

तकनीकी विशिष्टियाँ /Technical Specifications

* जेम केटेगरी विशिष्टि के अनुसार / As per GeM Category Specification

विवरण/Specification	विशिष्टि का नाम /Specification Name	बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)
Standards	Conformity to Standards	CE (with 4 digit notified number), USFDA, BIS (As per medical directive)
Performance Parameters	Capacity in liters	851-900 L
	Type of Cabinet	Vertical
	Should have heavy duty lockable castors and levelling for adjustments and installation	Yes
	Number of holding racks	15, 25, 24, 30, 20, 0, 1
	Number of inner storage compartments with insulated doors and adjustable height	3, 4, 5, 1
	Number of compressor	1, 2
	Warranty of Freezer (except compressor) in Years (from the date of installation)	3, 4, 5, 1
Additional Parameters	Number of corrosion resistant racks	2, 3, 4, 5, 0
	Pairs of Cryogloves to be provided	1, 2
	Number of Ice scraper to be provided	1, 2, 3
	Tempearture Chart recorder to be provided	Yes, No
	Availability of USB port to download event log and freezer setting	Yes

परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.No	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Heera Ram	342011,Residency Road, Ratanada, Jodhpur	2	65

Special terms and conditions-Version:1 effective from 11-11-2022 for category Ultra low temperature laboratory deep freezer

1.

1. **Comprehensive warranty:** Comprehensive warranty shall include preventive maintenance including calibration as per technical/ service /operational manual of the manufacturer, service charges and spare parts. During the warranty period commencing from date of the successful completion of warranty period, Seller's personnel shall visit each consignee site as recommended in the manufacturer's technical/ service /operational manual, at least once in six months. warranty shall not be including the consumables. Further there will be 98% uptime warranty during warranty period on 24 (hrs) X 7 (days) X 365 (days) basis, with a penalty, to extend warranty period by double the downtime period.
2. **Service centres:** Details of Service outlets in India to render services for equipment to be furnished to buyer/consignees with complete address, telephone numbers, e mails etc at time of making the supplies shall be the responsibility of seller to ensure that authorized service centres are available to cater to the areas where supplies are made within reasonable distance from where the service calls can be handled. Details of toll-free numbers for service call and online registration of service requests also to be provided buyer/consignee at the time of supplies.
3. **Source of supply:** It shall be responsibility of seller to provide Documents regarding source of equipment such as copy of Performance invoice or any other documents to establish that the products supplied are manufactured by OEM indicated and sourced from them.
4. **Packing and Marking:** Medical equipments being very delicate and sensitive packing for the goods shall be strong and durable enough to withstand transit including transshipment (if any), rough handling, open storage etc. without any damage, deterioration etc. .The size, weights and volumes of the packing case, remoteness of the final destination of the goods, availability or otherwise of transport and handling facilities at all points during transit up to final destination,. Quality of packing, the manner of marking within & outside the packages and provision of accompanying documentation shall take into consideration the nature of medical equipments being supplied. The accessories shall be suitably labelled and packed. Each of the package shall be marked on three sides with indelible paint of proper quality: indicating contract number and date, brief description of goods including quantity, Packing list reference number, country of origin of goods and any other relevant details.
5. **Spare Parts:** Seller shall provide materials, information etc. pertaining to spare parts manufactured and supplied by the OEM. It shall be ensured that the required spares are available for purchase at least for 5 years from date of supplies. In case due to any reasons the production of the spare parts is discontinued sufficient advance notice should be given to the buyer/consignee before such discontinuation to provide adequate time to purchase the required spare parts etc. Further, OEM and their service centres/dealers shall carry sufficient inventories to assure ex-stock supply of consumables and spares for the equipment so that the same are available. OEM or reseller shall always accord most favoured client status to the buyer/consignee and shall give the most competitive prices for spares and consumables of its machines/equipments supplied.
6. **Installation, Training, Manuals:** Seller shall be responsible to carry out Installation & commissioning, Supervision and Demonstration of the goods. They shall provide required jigs and tools for assembly, and civil works for the completion of the installation and Training of Consignee's representatives for operation and maintaining the equipment and supplying required number of operation & maintenance manual for the goods. In case the category parameters are specifying any requirements regarding the installations, training and manuals the same shall also be applicable.
7. **Electrical safety checking:** Sellers are required to make sure that they furnish the list of equipments carrying out routine and preventive maintenance to buyer/consignee .They should make sure to periodically check the electrical safety aspects as per BIS Safety Standards or equivalent .In case they do not have required equipment for such testing should ensure that the equipments checked for electrical safety compliance through labs with facilities for such checking during every preventive maintenance cycle.
8. **Software:** All software updates should be provided free of cost during warranty period.
9. Any other Terms and Conditions which is not included or at variance with the conditions specified in the Bid, may be added by the buyer through Additional Terms and Conditions in the bid to ensure that they are procured from authentic/validated source with appropriate and applicable quality. The above terms and conditions are in reverse order of precedence i.e. ATC shall supersede specific STC which shall supersede General Terms and Conditions ("GTC"), whenever there are any conflicting provisions.

क्रेता द्वारा जोड़ी गई बिड की विशेष शर्तें/Buyer Added Bid Specific Terms and Conditions

1. Generic

OPTION CLAUSE: The Purchaser reserves the right to increase or decrease the quantity to be ordered up to 25 percent of bid quantity at the time of placement of contract. The purchaser also reserves the right to increase or decrease the ordered quantity up to 25% of the contracted quantity during the currency of the contract at the contracted

rates. The delivery period of quantity shall commence from the last date of original delivery order and in case where option clause is exercised during the extended delivery period the additional time shall commence from the last date of extended delivery period. The additional delivery time shall be $(\text{Increased quantity} \div \text{Original quantity}) \times \text{Original delivery period (in days)}$, subject to minimum of 30 days. If the original delivery period is than 30 days, the additional time equals the original delivery period. The Purchaser may extend this calculate delivery duration up to the original delivery period while exercising the option clause. Bidders must comply with these terms.

2. Buyer Added Bid Specific ATC

Buyer Added text based ATC clauses

Required Specifications

S. N	Technical Specification	
	Specification	Point Wise
1	Capacity and display	Freezer should be of 850-900 Liters capacity and should have ar and legible eye-level advanced LED display
2	Temperature monitoring and data export	System should have on board Temperature monitoring and data export via USB for documentation; with memory to provide on board data storage for at least a month or longer
3	Backup system	System should have option to connect to third part data monitoring and Dedicated alarm and backup system for 24/7 sample safety
4	Programmable operating temperature	System should have Programmable operating temperature from - 50°C up to 86°C with 1°C increment at 32°C Maximum ambient operating temperature.
5	Energy efficient	System should be highly energy efficient, with energy consumption around 11.8 KWh/day (Kilo Watt Hours per day) when system is at - 80°C [8.1KWh/day when system is at - 70°C]
6	Programmable microprocessor	Fully programmable microprocessor controlled with membranepad and eye level control panel.
7	Insulation	Insulation should be a combination of advanced vacuum insulation panel (VIP) technology & Polyurethane foam to reduce free wall thickness, while maintaining highest heat insulation. This provides up to 30% more storage capacity than traditionally insulated freezers without increasing the external dimensions.
8	System Exterior and interior	System Exterior should be made up of powder coated Steel to resist scratch and rust and the interior should be of Polished Stainless-Steel grade 304 2B for easy cleaning and to eliminate potential for oxidation.
9	Inner doors	At least 3 Inner doors should have silicone seal to prevent temperature loss and Outer door should have safe High efficiency seal

10	Compartments	At least 3 inner compartments should be available.
11	Temperature pull down timing	Ambient (20 °C +/-1) to -85°C pull down timing should be 4 h 10 min or lesser; with freezer being maintained empty
12	Warm up time	Warm up time (freezer 2/3 full, from -85 °C to 0 °C) of at least hrs. or longer
13	Heavy duty castors wheels	System should have Heavy duty castors wheels to easily move e freezer to a new position
14	Cabinet uniformity	Freezer should have Cabinet uniformity (freezer set to -80 °C) +4.3°C/-0.9°C or better with Enhanced shelf design which has egrated airways incorporated for efficient and consistent air circulation in the chamber and rounded metal rims of shelf edg for enhanced safety of users' fingers.
15	Automated vent port	Freezers should have an Automated vent port for quick re-acc to samples and front panel air filter, easily accessible for clear g and maintenance.
16	Noise level	System should have Silence level/Noise level (at-80 °C) 47.8 d r lesser
17	Security keyed locks	Should have security keyed locks on the outer doors and lids k p out unauthorized users.
18	Battery back-up, alarm, securit y	Freezer must have battery back-up for display settings and alm in case of power outage, Password protection security to pr ent unauthorized tampering of freezer settings.
19	Remote alarm port	Freezer must have Remote alarm port and RS 485 interface de logging port
20	HFC-free. CFC-FREE, HCFC-FREE non-flammable refrigerants	Freezer must have HFC-free, CFC-FREE, HCFC-FREE non-flammr le refrigerants, and should have Hydrocarbon (HC)/ Natural Ga ased refrigeration system. It must be energy efficient and herr tically sealed two stage cascade refrigeration system
21	ISO 9001 standard quality test	Freezer must have ISO 9001 standard quality test requirement and IEC 61010 Electrical safety CE certified.
22	Capacity	Freezer must have capacity to hold 18 racks and 576 (57600 s mples) boxes of 2" height vials in the 3-inner compartment va nt Freezer should have electric supply of 230v/50hz, 10 amps. ngle phase.
23	Stabilizer	Freezer must supply with Branded 5 KVA Stabilizer
24	Warranty	Three years

3. Buyer Added Bid Specific ATC

Buyer uploaded ATC document [Click here to view the file.](#)

अस्वीकरण/Disclaimer

The additional terms and conditions have been incorporated by the Buyer after approval of the Competent Authority Buyer Organization, whereby Buyer organization is solely responsible for the impact of these clauses on the bidding process, its outcome, and consequences thereof including any eccentricity / restriction arising in the bidding process due to these ATCs and due to modification of technical specifications and / or terms and conditions governing the bid. Any clause(s) is / are incorporated by the Buyer regarding following, the bid and resultant contracts shall be treated null and void and such bids may be cancelled by GeM at any stage of bidding process without any notice:-

1. Definition of Class I and Class II suppliers in the bid not in line with the extant Order / Office Memorandum issued by DPIIT in this regard.
2. Seeking EMD submission from bidder(s), including via Additional Terms & Conditions, in contravention to exemption provided to such sellers under GeM GTC.
3. Publishing Custom / BOQ bids for items for which regular GeM categories are available without any Category being bunched with it.
4. Creating BoQ bid for single item.
5. Mentioning specific Brand or Make or Model or Manufacturer or Dealer name.
6. Mandating submission of documents in physical form as a pre-requisite to qualify bidders.
7. Floating / creation of work contracts as Custom Bids in Services.
8. Seeking sample with bid or approval of samples during bid evaluation process. (However, in bids for [attached categories](#), trials are allowed as per approved procurement policy of the buyer nodal Ministries)
9. Mandating foreign / international certifications even in case of existence of Indian Standards without specifying equivalent Indian Certification / standards.
10. Seeking experience from specific organization / department / institute only or from foreign / export experience.
11. Creating bid for items from irrelevant categories.
12. Incorporating any clause against the MSME policy and Preference to Make in India Policy.
13. Reference of conditions published on any external site or reference to external documents/clauses.
14. Asking for any Tender fee / Bid Participation fee / Auction fee in case of Bids / Forward Auction, as the case may be.
15. Buyer added ATC Clauses which are in contravention of clauses defined by buyer in system generated bid template as indicated above in the Bid Details section, EMD Detail, ePBG Detail and MII and MSE Purchase Preference sections of the bid, unless otherwise allowed by GeM GTC.
16. In a category based bid, adding additional items, through buyer added additional scope of work/ additional terms and conditions/or any other document. If buyer needs more items along with the main item, the same must be added through bunching category based items or by bunching custom catalogs or bunching a BoQ with the main category based item, the same must not be done through ATC or Scope of Work.

Further, if any seller has any objection/grievance against these additional clauses or otherwise on any aspect of this they can raise their representation against the same by using the Representation window provided in the bid details in Seller dashboard after logging in as a seller within 4 days of bid publication on GeM. Buyer is duty bound to reply to such representations and would not be allowed to open bids if he fails to reply to such representations.

All GeM Sellers/Service Providers shall ensure full compliance with all applicable labour laws, including provisions, rules, schemes and guidelines under the four Labour Codes i.e. the Code on Wages, 2019; the Industrial Relations Code, 2020; the Occupational Safety, Health and Working Conditions Code, 2020; and the Code on Social Security, 2020 as and when notified and brought into force by the Government of India.

For all provisions of the Labour Codes that are pending operationalisation through rules, schemes or notifications, the corresponding provisions of the pre-existing labour enactments (such as The Minimum Wages Act, 1948, The Payment of Wages Act, 1936, The Payment of Bonus Act, 1965, The Equal Remuneration Act, 1976, The Payment of Gratuity Act, 1972, etc. and relevant State Rules) shall continue to apply.

to remain applicable.

The Seller/ Service Providers shall, therefore, be responsible for ensuring compliance under:

- **All notified and enforceable provisions of the new Labour Codes as mentioned hereinabove; and**
- **All operative provisions of the erstwhile Labour Laws until their complete substitution.**

All obligations relating to wages, social security, safety, working conditions, industrial relations etc. an any other statutory requirements shall be strictly met by the Seller/ Service Provider. Any non-compliance shall constitute a breach of the contract and shall entitle the Buyer to take appropriate action in accordance with the contract and applicable law.

[यह बिड सामान्य शर्तों के अंतर्गत भी शासित है /This Bid is also governed by the General Terms and Conditions](#)

जेम की सामान्य शर्तों के खंड 26 के संदर्भ में भारत के साथ भूमि सीमा साझा करने वाले देश के बिडर से खरीद पर प्रतिबंध के संबंध में भा: साथ भूमि सीमा साझा करने वाले देश का कोई भी बिडर इस निविदा में बिड देने के लिए तभी पात्र होगा जब वह बिड देने वाला सक्षम प्राधिक पास पंजीकृत हो। बिड में भाग लेते समय बिडर को इसका अनुपालन करना होगा और कोई भी गलत घोषणा किए जाने व इसका अनुपालन न पर अनुबंध को तत्काल समाप्त करने और कानून के अनुसार आगे की कानूनी कार्रवाई का आधार होगा।/In terms of GeM GTC clause 26 regarding Restr on procurement from a bidder of a country which shares a land border with India, any bidder from a country which shares a land border with India will be eligible to bid in this only if the bidder is registered with the Competent Authority. While participating in bid, Bidder has to undertake compliance of this and any false declaration and non-compliance this would be a ground for immediate termination of the contract and further legal action in accordance with the laws.

---धन्यवाद/Thank You---